Enclosure (variant)	Sensor minus air (°C)		RH error (points)
	calm	5 m s^{-1}	calm
Baseline closed box (V0)	−4.028	−1.390	+13.24
Passive multi-plate shield (V1)	−0.012	—	+0.03
Actively aspirated reference (V2)	−0.005	—	+0.01