Training Checklist: **Abbott HIV Combo**

INSTRUCTIONS: The Trainee will initial next to each Training Task Description upon reviewing the contents of all checklist items. The Trainer will initial next to each Training Task Description upon training the Trainee. Upon successful completion of this training checklist, the Trainee and Trainer will sign and date the checklist at the end.

Training Task Description	Trainee Initials	Trainer Initials
External Quality Control (QC) a) External QC is tested during training and with each new lot/shipment. b) The following data must be recorded when QC is run. QC level, QC lot number, QC expiration date, QC results, test date, and testing personnel initials. c) There are 4 levels of QC that are run, a Non-reactive control, HIV-1 reactive control, HIV-2 reactive control, and HIV-1 p24 Antigen reactive control.		
Test Strip Handling a) Unused test strips must be stored sealed in the aluminum pouch with the included desiccant properly at all times. b) If the desiccant package is missing from the aluminum pouch, DO NOT USE, discard all test cards in the pouch and open a new pouch. c) Test strips must not be expired prior to use. d) Test strips must be used within 2 hours after removing the protective foil cover from each strip. e) Do not touch the sample pad with your fingers.		
Sample Handling a) Positively identify the patient's sample prior to testing. b) Capillary fingerstick or venous whole blood collected in EDTA may be used. c) Capillary fingerstick blood specimens must be tested immediately. d) Capillary fingerstick samples are obtained from the second drop of blood with the capillary tube. e) Venous whole blood specimens may be stored 2 days at ambient temperatures prior to testing.		
Test Procedure a) When ready for testing, remove the protective foil from the test strip. b) Apply 50 uL (microliters) of sample by touching the tip of the pipette to the Sample Pad. c) After sample application, wait 1 minute. d) Add 1 drop of Chase Buffer to the Sample Pad. e) Read the test result between 20-30 minutes after addition of the Chase Buffer. (No more than 30 minutes)		
Results Interpretation a) A Control line should always appear in the Control Area. If no control line is present, the results are invalid and the test should be repeated with a new test Cassette. b) Any pink or red lines showing in the Ag or Ab test areas in conjunction with the presence of a control line are interpreted as Preliminary Positive.		
Support a) Know to call Alere Technical Support when requiring technical assistance, troubleshooting, and error reports – 1 877 441 7440 or 321 441 720.		
Trainee to run <u>one</u> level of QC and enter results into the HIV QC Log below.		

Training Checklist: **Abbott HIV Combo**

QC LOG



Date	Control Lot # / Exp	Strip Lot # / Exp	Non-reactive ctrl	HIV-1 ctrl	HIV-2 ctrl	HIV-1 p24 Ag ctrl	Pass? (Y/N)	Initials
1/1/24	1234-567-89101 1/30/25	123456 1/1/26	NR	R	R	R	Y	JP

Trainee signature:		Date:	
Trainer signature:		Date:	