BIOFIRE® SPOTFIRE® RESPIRATORY/SORE THROAT (R/ST) PANEL TRAINING CHECKLIST

Trainee Name: _____

Trainer Name: _____

Date: _____

Trainee has reviewed the SPOTFIRE R/ST Panel Instructions for Use (IFU) and SPOTFIRE R/ST Panel Quick Guide prior to performing testing.	**Trainee Initials:** _____
Operator trainee has performed the Positive External Quality Control run and received the expected result.	**Trainee Initials:** _____
Operator trainee has performed the Negative External Quality Control run and received the expected result.	**Trainee Initials:** _____

Scan for a Digital Copy

BFR0002-8617-01

BIOFIRE® SPOTFIRE® System and BIOFIRE® SPOTFIRE® Respiratory/Sore Throat (R/ST) Panel Quiz

Operator Name: _____

Date: _____

This quiz is designed as an educational tool only and is not to be used as sole evidence of operator competency to run the SPOTFIRE R/ST on the SPOTFIRE System. Always follow manufacturer instructions in the Instructions for Use (IFU) and Operator's Manual documents.

1. **At what temperature should the SPOTFIRE R/ST Panel pouches be stored?**
   - A. Room Temperature (15° to 25° C)
   - B. Frozen (-20° to 0° C)
   - C. Refrigerated (0° to 2° C)
   - D. Frozen (-20° to 0° C) or Refrigerated (0° to 2° C)

2. **Which are the approved sample type(s) for use with the SPOTFIRE R/ST Panel?**
   - A. Nasopharyngeal swab in transport media
   - B. Nasal swab in transport media or saline
   - C. Throat swab in transport media
   - D. A & C

3. **Due to the sensitive nature of the SPOTFIRE R/ST Panel, which of the following guidelines for contamination prevention should be followed?**
   - A. Do not prepare specimens or load pouches in areas used to prepare vaccine material and/or perform viral and bacterial culture
   - B. Handle specimens and pouches one at a time
   - C. Change gloves and clean the work area between each patient specimen
   - D. All of the above

4. **After the sample injection vial has been loaded with patient sample and buffer solution, the injection vial should be capped and mixed by:**
   - A. Shaking vigorously
   - B. Gently inverting 3 times
   - C. Centrifugation
   - D. Tapping the base of the injection vial on the counter

5. **What is the appropriate cleaning solution for the loading block and work surface where SPOTFIRE R/ST Panel samples are prepared?**
   - **A.** Quaternary ammonium solution
   - **B.** Isopropyl alcohol solution
   - **C.** 10% bleach solution followed by a water wipe
   - **D.** Hydrogen peroxide solution

6. **The sample buffer is harmful if swallowed and can cause serious eye damage and/or skin irritation.**
   - **A.** True
   - **B.** False

7. **The sample preparation work area should always be cleaned between EVERY patient sample.**
   - **A.** True
   - **B.** False

8. **At what temperature should the SPOTFIRE R/ST Panel External Control Material be stored?**
   - **A.** Refrigerated to Room Temperature (2° to 25° C)
   - **B.** Frozen (-20° to 0° C)
   - **C.** Refrigerated (0° to 2° C)
   - **D.** Frozen (-20° to 0° C) or Refrigerated (0° to 2° C)

9. **The SPOTFIRE R/ST Panel External Control Testing needs to be performed when?**
   - **A.** When training a new operator
   - **B.** Every new lot or shipment of SPOTFIRE reagent pouches is received
   - **C.** A monthly negative QC test should be ran at each testing location to monitor environmental contamination
   - **D.** All of the above

10. **Routine QC should be run by _____.**
   - **A.** Loading the pouch as usual
   - **B.** Going into QC mode and selecting the appropriate QC type being run
   - **C.** Both A & B
   - **D.** None of the above

11. **The manufacturer recommends restarting instrument how frequently?**
   - **A.** Never; it is ok to stay on indefinitely
   - **B.** Once a day
   - **C.** Once a week
   - **D.** Once a month

BFR0002-2377-01

BIOFIRE® SPOTFIRE® System and BIOFIRE® SPOTFIRE® Respiratory/Sore Throat (R/ST) Panel Quiz Key

Operator Name: _____

Date: _____

This quiz is designed as an educational tool only and is not to be used as sole evidence of operator competency to run the SPOTFIRE R/ST on the SPOTFIRE System. Always follow manufacturer instructions in the Instructions for Use (IFU) and Operator's Manual documents.

1. **At what temperature should the SPOTFIRE R/ST Panel pouches be stored?**
   - A. **Room Temperature (15° to 25° C)**
   - B. Frozen (-20° to 0° C)
   - C. Refrigerated (0° to 2° C)
   - D. Frozen (-20° to 0° C) or Refrigerated (0° to 2° C)

2. **Which are the approved sample type(s) for use with the SPOTFIRE R/ST Panel?**
   - A. Nasopharyngeal swab in transport media
   - B. Nasal swab in transport media or saline
   - C. Throat swab in transport media
   - D. **A & C**

3. **Due to the sensitive nature of the SPOTFIRE R/ST Panel, which of the following guidelines for contamination prevention should be followed?**
   - A. Do not prepare specimens or load pouches in areas used to prepare vaccine material and/or perform viral and bacterial culture
   - B. Handle specimens and pouches one at a time
   - C. Change gloves and clean the work area between each patient specimen
   - D. **All of the above**

4. **After the sample injection vial has been loaded with patient sample and buffer solution, the injection vial should be capped and mixed by:**
   - A. Shaking vigorously
   - B. **Gently inverting 3 times**
   - C. Centrifugation
   - D. Tapping the base of the injection vial on the counter

5. **What is the appropriate cleaning solution for the loading block and work surface where SPOTFIRE® R/ST Panel samples are prepared?**
   - A. Quaternary ammonium solution
   - B. Isopropyl alcohol solution
   - C. **10% bleach solution followed by a water wipe**
   - D. Hydrogen peroxide solution

6. **The sample buffer is harmful if swallowed and can cause serious eye damage and/or skin irritation.**
   - A. **True**
   - B. False

7. **The sample preparation work area should always be cleaned between EVERY patient sample.**
   - A. **True**
   - B. False

8. **At what temperature should the SPOTFIRE R/ST Panel External Control Material be stored?**
   - A. **Refrigerated to Room Temperature (2° to 25° C)**
   - B. Frozen (-20° to 0° C)
   - C. Refrigerated (0° to 2° C)
   - D. Frozen (-20° to 0° C) or Refrigerated (0° to 2° C)

9. **The SPOTFIRE R/ST Panel External Control Testing needs to be performed when?**
   - A. When training a new operator
   - B. Every new lot or shipment of the SPOTFIRE reagent pouches is received
   - C. A monthly negative QC test should be run at each testing location to monitor environmental contamination
   - D. **All of the above**

10. **Routine QC should be run by _____.**
   - A. Loading the pouch as usual
   - B. Going into QC mode and selecting the appropriate QC type being run
   - C. **Both A & B**
   - D. None of the above

11. **The manufacturer recommends restarting instrument how frequently?**
   - A. Never; it is ok to stay on indefinitely
   - B. Once a day
   - C. **Once a week**
   - D. Once a month

BFR0002-2376-01