Training Checklist: **Clinitek Status Urine hCG**

INSTRUCTIONS: The Trainee will initial next to each Training Task Description upon reviewing the contents of all checklist items. The Trainer will initial next to each Training Task Description upon training the Trainee. Upon successful completion of this training checklist, the Trainee and Trainer will sign and date the checklist at the end.

Training Task Description	Trainee Initials	Trainer Initials
External Quality Control (QC) a) External QC is tested with each new lot/shipment and monthly. b) The following data must be recorded when QC is run. QC for test type, QC level, QC lot number, QC expiration date, QC results, test date, and testing personnel initials.		
Maintenance a) Know to complete daily and weekly preventive maintenance procedures. **DAILY** Decontaminate display and analyzer surfaces **WEEKLY** Clean test table and table insert Clean the white calibration bar		
Cassette Handling a) Cassettes are stored at room temperature or refrigerated. b) Cassettes must be brought to room temperature before opening and testing. c) Do not use cassettes beyond the expiration date.		
Sample Handling a) Positively identify the patient and label the urine sample. b) The urine sample should be tested within 2 hours of sample collection. c) If testing is delayed >2 hours, urine specimens may be stored in the refrigerator up to 72 hours prior to testing. The sample must come to room temperature prior to testing. d) Unacceptable urine specimens include urine specimens with preservatives.		
Test Procedure a) Select “Cassette Test,” then “CLINITEST hCG cassette.” b) Position test table insert is facing up and in position for a cassette test. c) Remove test cassette from package and place on test table. d) Touch start; operator has 8 seconds to add urine sample into the well on the cassette. e) Test table and cassette will automatically be pulled into the instrument.		
Results a) Results will display after analysis is complete. b) Invalid results will be automatically determined by the instrument.		
Support a) Know to call Siemens Healthcare Technical Support when requiring technical assistance, troubleshooting, and error reports – 1 877 229 3711.		

Trainee signature:		Date:	
Trainer signature:		Date:	