Training Checklist: **Coag-Sense**

INSTRUCTIONS: The Trainee will initial next to each Training Task Description upon reviewing the contents of all checklist items. The Trainer will initial next to each Training Task Description upon training the Trainee. Upon successful completion of this training checklist, the Trainee and Trainer will sign and date the checklist at the end.

Training Task Description	Trainee Initials	Trainer Initials
Analyzer Operation a) Know how to turn on and off the analyzer. b) Know to verify self-check passes when analyzer is powered on.		
External Quality Control (QC) c) External QC is tested during training and with each new lot/shipment. d) The following data must be recorded when QC is run. QC for test type, QC level, QC lot number, QC expiration date, QC results, test date, and testing personnel initials. e) There are low control strips and high control strips shipped with each strip kit. f) Prior to QC testing, make sure the correct strip lot information is displayed for the strips provided in the kit. g) Control activation solution is dispensed into a glass capillary tube and is applied to the control strip when prompted by the instrument. h) Ensure all control results correspond with the expected results.		
Maintenance a) Know to complete daily preventive maintenance procedures. **DAILY** ☐ Clean and disinfect the instrument		
Sample Handling and Test Procedure a) Select “Test” on the home screen for patient testing. b) Confirm the patient strip lot info on the kit matches the info on the analyzer. c) Enter the correct patient identifier into the analyzer. d) Gently push patient test strip into the meter; the meter warm up in 25 seconds. e) Do not lance finger until meter displays “APPLY SAMPLE.” (sample must be applied to the strip within 2.5 minutes) f) Know to touch the tip of the sample transfer tube to the bead of blood and let capillary action fill until blood flow stops at the white plug. g) Immediately place patient blood from sample transfer tube into the test strip. h) If “Testing Please Wait” message does not display within 8 seconds, not enough blood was added. Repeat test procedure with a new strip and fingerstick.		
Results Interpretation a) The results will display on the analyzer’s screen. b) If the meter displays a message instead of results, refer to the “Troubleshooting” section of the manual for assistance.		
Support a) Know to call Coag-Sense Technical Support when requiring technical assistance, troubleshooting, and error reports – 1 866 903 0890 or Techsupport@coagusense.com		

Trainee signature:		Date:	
Trainer signature:		Date:	