# Training Checklist: **Mckesson Drugs of Abuse Test Cup**

INSTRUCTIONS: The Trainee will initial next to each Training Task Description upon reviewing the contents of all checklist items. The Trainer will initial next to each Training Task Description upon training the Trainee. Upon successful completion of this training checklist, the Trainee and Trainer will sign and date the checklist at the end.

| Training Task Description | Trainee Initials | Trainer Initials |
|---|---|---|
| **Internal Quality Control (QC)**<br>a) Know how to check for internal procedural control with each patient test.<br>b) The internal procedural control confirms sufficient specimen volume, adequate membrane wicking, and correct procedural technique. | | |
| **Reagent Handling**<br>a) The Test Device must remain in the sealed pouch until use.<br>b) The Test Device is stored at room temperature or refrigerated temperatures away from direct sunlight, moisture, and heat.<br>c) Do not use beyond the expiration date. | | |
| **Sample Handling**<br>a) Patient should collect the urine sample directly into the Test Cup. | | |
| **Test Procedure**<br>a) Allow the Test Cup to come to room temperature prior to testing.<br>b) Label the Test Cup with patient's information.<br>c) Test Cup should be filled with urine between 30-110 mL as marked on the cup. Secure cap firmly onto the cup.<br>d) Check urine temperature at 2-4 minutes by reading green color on the temperature indicator strip.<br>e) Remove peel-off label; read strip pads between 2-5 minutes. (Do not read results after 10 minutes) | | |
| **Results Interpretation**<br>a) Negative results: two lines present – one in the control region, one in the test region. A faint pink line in the test region is still considered negative.<br>b) Positive results: one red line present – no line appears in the test region. | | |
| **Support**<br>a) Know to call Technical Support when requiring technical assistance, troubleshooting, and errors – 1 866 982 3818. | | |

| | | | |
|---|---|---|---|
| Trainee signature: | | Date: | |
| Trainer signature: | | Date: | |