Training Checklist: **Piccolo**

INSTRUCTIONS: The Trainee will initial next to each Training Task Description upon reviewing the contents of all checklist items. The Trainer will initial next to each Training Task Description upon training the Trainee. Upon successful completion of this training checklist, the Trainee and Trainer will sign and date the checklist at the end.

Training Task Description	Trainee Initials	Trainer Initials
Analyzer Operation a) Know general Analyzer menu and use. b) Know how to recall and print records. c) Know how to install the printer paper roll: a. Locate the paper roll compartment situated on top of the Piccolo. b. Pull the cover up and away from you following the arrow marked on top. c. Place the paper roll into the compartment with the white side facing you. d. Snap down the cover.		
Maintenance a) Know to complete weekly, monthly, and bi-annual preventive maintenance procedures. **WEEKLY** Clean analyzer external case and display **MONTHLY** ***Perform external Controls** **BI-ANNUALLY (April, October)** Clean air filter in the back of the analyzer		
External Quality Control (QC) a) External QC is tested at least every 30 days, whenever laboratory conditions have changed significantly, training, and each new lot/shipment. b) The following data must be recorded when QC is run. QC on disc type, QC level, QC lot number, QC expiration date, QC results, test date, and testing personnel initials. c) There are 2 levels of QC that are run, a high and a low control. d) Prior to QC testing, thaw controls at room temperature for 1 hour before using. Mix the QC vial gently by using a windshield wiper motion several times, then upright the vial and gently swirly it several times, then tap the top of the vial a few times. e) Do not refreeze the control samples after use. Discard used controls appropriately. f) Ensure all control results are within the product insert values for the control lot number you are using. g) If control results are not within the expected value range, repeat the testing.		
Disc Handling a) Do not allow discs sealed in their foil pouches to remain at room temperature for longer than 48 hours prior to use. b) Discs are meant to be used cold and immediately after removing from the refrigerator. c) After a foil pouch has been opened, the reagent disc must be used within 20 minutes or discarded. d) Do not use a disc from a damaged pouch. e) Do not use a reagent disc that has been dropped. f) Identify where the sample port is on the disc.		

Training Checklist: **Piccolo**

Sample Handling a) Whole blood samples collected into the lithium heparin green tube is used for the test. b) Whole blood venipuncture samples should be run within 60 minutes of collection. c) Whole blood sample must be inverted several times just prior to sample transfer into the reagent disc. d) Wipe off any residue of sample left on top of the disc with a lint-free wipe. e) Once the sample is placed into the disc, the test should be run within 10 minutes.		
Test Procedure a) Analysis must begin immediately, no more than 10 minutes, after dispensing the sample into the reagent disc. b) Use a pipette to draw up the sample (or control). c) Hold the reagent disc level and place the pipettor into the sample port at a 45 degree angle. d) Slowly dispense the sample into the sample port, ensuring no air bubbles are introduced into the sample port. Air bubbles can be visually seen if introduced. e) Keep the prepared disc flat. f) On the instrument, press ANALYZE on the touch screen to open the instrument drawer.		
Maintenance a) Know to clean the analyzer exterior. b) Know to clean the air filter according to the Preventive Maintenance logs.		
Results Interpretation a) Review whenever test results do not match the patient's symptoms or clinical findings. b) Any result for a particular test that exceeds the assay range should be sent to a referral laboratory. c) Review the Results Interpretation page at the end of this training to understand the possible results that require intervention prior to releasing the results for patient use.		
Support a) Know to call Piccolo Technical Support when requiring technical assistance, troubleshooting, and error reports – 1 800 822 2947.		
Trainee to run <u>one</u> level of Piccolo QC for any disc and attach results to the Piccolo QC Log below.		

Training Checklist: **Piccolo**

QC LOG



Piccolo Info

Disc Type (circle one): BMP | CMP | Lipid | Liver | Metlyte | CRP | MetLac 12

Disc Exp: _____ / _____

Ctrl 1 Lot#/Exp: _____ / _____

Ctrl 2 Lot#/Exp: _____ / _____

General Info

QC Test Date: _____

Testing Personnel Initials: _____

LEVEL 1 CONTROL	LEVEL 2 CONTROL
(Attach Results Here)	(Attach Results Here)

Training Checklist: **Piccolo**

Results Interpretations

~ ~ ~ (Tildes)

A suppressed test result.

Test cannot be determined due to any of the following:

- sample bead mix problem
- nonlinear reaction
- endpoint reaction not reached
- concentration outside the analyzers capability

Test Result

piccolo xpress
Comprehensive Metabolic
14 Oct 2009 07:29 AM
Sample Type: Patient
Patient ID: 70439-0
Gender: Male
Race: White
Operator ID: 1
Disc Lot Number: 9411BC4
Serial Number: 0000P03015

NA+	130	128-145	MMOL/L
K+	3.9	3.6-5.1	MMOL/L
tCO2	~~~	18-33	MMOL/L
CL-	101	98-108	MMOL/L
GLU	128 *	73-118	MG/DL
CA	9.6	8.0-10.3	MG/DL
BUN	~~~	7-22	MG/DL
CRE	1.5 *	0.6-1.2	MG/DL
ALP	42	53-128	U/L
ALT	17	10-47	U/L
AST	22	11-38	U/L
TBIL	0.7	0.2-1.6	MG/DL
ALB	3.8	3.3-5.5	G/DL
TP	7.6	6.4-8.1	G/DL

QC OK
(HEM 0 LIP 0 ICT 0)

Error Report

piccolo xpress
Comprehensive Metabolic
14 Oct 2009 07:29 AM
Sample Type: Patient
Patient ID: 70439-0
Gender: Male
Race: White
Operator ID: 1
Disc Lot Number: 9411BC4
Serial Number: 0000P03015

01	EDEE	FFFF	6CFF	FFFF
02	B9DF	FFFF	9940	FFFF
03	0000	0000	0000	0000
04	0000	0000	0000	0000
05	0000	0000	0000	0000
06	0000			

NA+	130	0000	0000	MMOL/L
K+	3.9	0000	0000	MMOL/L
tCO2	~~~	2000	0000	MMOL/L
CL-	101	0000	0000	MMOL/L
GLU	128 *	0000	0000	MG/DL
CA	9.6	0000	0000	MG/DL
BUN	~~~	0020	0000	MG/DL
CRE	1.5 *	0000	0000	MG/DL
ALP	42	0000	0000	U/L
AST	22	0000	0000	U/L
ALT	17	0000	0000	U/L
TBIL	0.7	0000	0000	MG/DL
ALB	3.8	0000	0000	G/DL
TP	7.6	0000	0000	G/DL

RQC: 100
(HEM -40 LIP 161 ICT 9)

Version: 2.1.33

Internal Troubleshooting Codes

*** (Asterisk)**

Test results are falling outside of the patient normal reference range set in the analyzer.

Reagent RQC

Value used to evaluate the viability of the reagents at the time of the test run.

Reagent QC

"QC OK" prints when the reagents used to detect exposure to extreme conditions such as temperature and humidity are within expected ranges.

Sample Indices (Qualitative Interference)

HEM – Hemolysis (rupturing of red blood cells)

LIP – Lipemia (fat in the blood)

ICT – Icterus (bilirubin in the blood)

0 = None **1+** = Slight
2+ = Moderate **3+** = Gross

Sample Indices (Quantitative Interference)

There is a specific limit of interference for each individual analyte.

Software Version

Training Checklist: **Piccolo**

Test Result

piccolo xpress
Comprehensive Metabolic
14 Oct 2009 07:29 AM
Sample Type: Patient
Patient ID: 70439-0
Gender: Male
Race: Unknown
Operator ID: 1
Disc Lot Number: 9411BC4
Serial Number: 0000P03015

NA+	~~~	!	128-145	MMOL/L
K+	< 1.5	!	3.6-5.1	MMOL/L
tCO2	11	!	18-33	MMOL/L
CL-	~~~	!	98-108	MMOL/L
GLU	62	!	73 -118	MG/DL
CA	5.6	!	8.0-10.3	MG/DL
BUN	15	!	7-22	MG/DL
CRE	0.7	!	0.6-1.2	MG/DL
eGFR	N/C			ML/MIN
ALP	38	!	53-128	U/L
ALT	18	!	10-47	U/L
AST	20	!	11-38	U/L
TBIL	11	!	18-33	MG/DL
ALB	2.5	!	3.3-5.5	G/DL
TP	3.7	!	3.6-5.1	G/DL

QC OK
HEM 0 LIP 0 ICT 0
! CONFIRM LOW RECOVERIES !

< or >

Test results have exceeded the dynamic range.

eGFR N/C

Estimated Glomerular Filtration Rate
(Not Calculated)

This appears when the operator has failed to enter one or more of the following when prompted by the analyzer:

- Age (≥ 18 and ≤ 70)
- Gender
- Race

In addition to the original test result, an error report may print out whenever the eGFR is not calculated.

! Confirm Low Recoveries !

! Indicates that multiple tests are lower than expected. A possible "short or diluted" sample.

Error Report

piccolo xpress
Comprehensive Metabolic
08 Dec 2010 03:27 PM
Sample Type: Patient
Patient ID: 71805 – 0
Disc Lot Number: 0353BA3
Serial Number: 0000P04302

01 B40F FFFF 6CFF FFFF
02 B9DF FFFF 5BAD B9EF
03 0000 0000 0000 0000
04 0000 0000 0000 0000
05 0000 0000 0000 0000
06 0000

RQC: 98
HEM ~~~ LIP ~~~ ICT ~~~

Version: 2.1.33
403D Sample Mix Error

4XXX

Disc Cancellation
4 Digit Error Code &
Error Message

Error Report

piccolo xpress
Lipid Panel
14 Jan 2011 01:27 PM
Sample Type: Patient
Patient ID: 33258 - 0
Disc Lot Number: 0353BA3
Serial Number: 0000P04302

01 B40F FFFF 6CFF FFFF
02 B9DF FFFF 5BAD B9EF
03 0000 0000 0000 0000
04 0000 0000 0000 0000
05 0000 0000 0000 0000
06 0000
CHOL 225 * 0000 0000 MG/DL
HDL LIP 0400 0000 MG//DL
TRIG 471 * 0000 0000 MG/DL
TC/H ~~~ 0000 0002
LDL LIP 0400 0000 MG/DL
VLDL LIP 0400 0000 MG/DL

RQC: 100
HEM 3 LIP 261 ICT 1

Version 2.1.33

LIP

(Lipemic Interference)

Lipemia has adversely affected the result of this test.

Trainee signature:		Date:	
Trainer signature:		Date:	