pocH-100*i*™ MAINTENANCE LOG

Daily Maintenance							Serial Number										Month							Year						
	1	2	3	4	5	6	7	8	9	10	11	12	13	14	15	16	17	18	19	20	21	22	23	24	25	26	27	28	29	30	31
Check Instrument Status																															
Perform Shutdown																															
Perform Quality Control																															
Initials																															

Bi-Weekly Maintenance or every 150 samples	date / initials	date / initials	date / initials
Clean Transducer			

Quarterly Maintenance or every 1500 samples	date / initials
Clean Waste Chamber	

As Needed Maintenance	date / initials	date / initials	date / initials
Dispose Waste Fluid			
Clean Sample Adaptor			
Replace Printer Paper			
Clean Aperture of TD Chamber			

Initial / check mark indicates value within instrument limits; maintenance procedure performed as required.

Reagent Replacement	lot # / date	lot # / date	lot # / date
pocH-pack D			
pocH-pack L			

Comments

Reviewed by Supervisor	
Name	
Date	

pocH-100*i*™ MAINTENANCE LOG

Daily Maintenance							Serial Number										Month							Year						
	1	2	3	4	5	6	7	8	9	10	11	12	13	14	15	16	17	18	19	20	21	22	23	24	25	26	27	28	29	30	31
Check Instrument Status																															
Perform Shutdown																															
Perform Quality Control																															
Initials																															

Bi-Weekly Maintenance or every 150 samples	date / initials	date / initials	date / initials
Clean Transducer			

Quarterly Maintenance or every 1500 samples	date / initials
Clean Waste Chamber	

As Needed Maintenance	date / initials	date / initials	date / initials
Dispose Waste Fluid			
Clean Sample Adaptor			
Replace Printer Paper			
Clean Aperture of TD Chamber			

Initial / check mark indicates value within instrument limits; maintenance procedure performed as required.

Reagent Replacement	lot # / date	lot # / date	lot # / date
pocH-pack D			
pocH-pack L			

Comments

Reviewed by Supervisor	
Name	
Date	