Training Checklist: **Quidel QuickVue Influenza A/B**

INSTRUCTIONS: The Trainee will initial next to each Training Task Description upon reviewing the contents of all checklist items. The Trainer will initial next to each Training Task Description upon training the Trainee. Upon successful completion of this training checklist, the Trainee and Trainer will sign and date the checklist at the end.

Training Task Description	**Trainee Initials**	**Trainer Initials**
External Quality Control (QC) a) External QC swabs are tested during with each new lot/shipment. b) The following data must be recorded when QC is run. QC swab type (positive or negative), QC lot number, QC expiration date, QC results, test date, and testing personnel initials. c) There are 2 control swabs that are run, a positive and a negative control.		
Test Kit Handling a) The test kits must remain at room temperature. b) Do not use expired test kits.		
Sample Handling a) Positively identify the patient and label the swab sample. b) Use the swabs supplied in the kit. c) Swabs should be tested as soon as possible after collection.		
Test Procedure a) Know to dispense the Reagent Solution vial (round bulb with twist top) into the Reagent Tube. b) Place patient swab sample into prepared Reagent Tube, roll swab 3x while pressing swab. Keep swab in tube for 1 minute. c) Roll the swab against the Reagent Tube as the swab is removed. d) Place Test Strip into Reagent Tube. e) Read results in 10 minutes. (No more than 10 minutes)		
Results Interpretation a) Any shade of pink-to-red Test Line (including faint pink lines) with the appearance of a blue procedural Control line indicates a positive result. b) A blue procedural Control Line and no pink Test Line is a presumptive negative result. c) If the blue procedural Control Line does not appear, even if any shade of pink-to-red Test Line appears, the result is invalid. A new test should be performed on a new swab and test strip.		
Support a) Know to call Quidel Technical Support when requiring technical assistance, troubleshooting, and error reports – 1 800 874 1517, option 5 or technicalsupport@quidel.com		
Trainee to run a <u>positive</u> QC swab and enter results into the Quidel Quickvue QC Log below.		
Trainee to run a <u>negative</u> QC swab and enter results into the Quidel Quickvue QC Log below.		

Training Checklist: **Quidel QuickVue Influenza A/B**

QC LOG



QuickVue

Date	Test Kit (Cov, Flu, RSV, Strep)	Swab Level	Control Swab Lot # / Exp	Kit Lot # / Exp	Results	Pass? (Y/N)	Initials
1/1/24	Flu	Pos	001234 1/30/25	002345 1/1/26	Pos	Y	JP
1/1/24	Flu	Neg	001235 1/30/25	002345 1/1/26	Neg	Y	JP

Trainee signature:		Date:	
Trainer signature:		Date:	