Training Checklist: **Quidel QuickVue SARS Antigen**

INSTRUCTIONS: The Trainee will initial next to each Training Task Description upon reviewing the contents of all checklist items. The Trainer will initial next to each Training Task Description upon training the Trainee. Upon successful completion of this training checklist, the Trainee and Trainer will sign and date the checklist at the end.

Training Task Description	Trainee Initials	Trainer Initials
External Quality Control (QC) a) External QC swabs are tested during with each new lot/shipment. b) The following data must be recorded when QC is run. QC swab type (positive or negative), QC lot number, QC expiration date, QC results, test date, and testing personnel initials. c) There are 2 control swabs that are run, a positive and a negative control.		
Test Kit Handling a) The test kits must remain at room temperature. b) Do not use expired test kits.		
Sample Handling a) Positively identify the patient and label the swab sample. b) Use the swabs supplied in the kit. c) Prior to collecting the swab, instruct patient to blow their nose. Take 15 seconds to collect the sample. d) Swabs should be tested as soon as possible after collection, up to 5 days at room temperature.		
Test Procedure a) Know to dispense the Reagent Solution vial (round bulb with twist top) into the Reagent Tube. b) Place patient swab sample into prepared Reagent Tube, roll swab 3x while pressing swab. Keep swab in tube for 1 minute. c) Roll the swab against the Reagent Tube as the swab is removed. d) Place Test Strip into Reagent Tube. e) Start the timer for 10 minutes. f) Read results in 10-15 minutes. (No more than 15 minutes)		
Results Interpretation a) Any shade of pink-to-red Test Line (including faint pink lines) with the appearance of a blue procedural Control line indicates a positive result. b) A blue procedural Control Line and no pink Test Line is a presumptive negative result. c) If the blue procedural Control Line does not appear, even if any shade of pink-to-red Test Line appears, the result is invalid. A new test should be performed on a new swab and test strip.		
Support a) Know to call Quidel Technical Support when requiring technical assistance, troubleshooting, and error reports – 1 800 874 1517, option 5 or technicalsupport@quidel.com		
Trainee to run a <u>positive</u> QC swab and enter results into the Quidel Quickvue QC Log below.		
Trainee to run a <u>negative</u> QC swab and enter results into the Quidel Quickvue QC Log below.		

Training Checklist: **Quidel QuickVue SARS Antigen**

QC LOG



QuickVue

Date	Test Kit (Cov, Flu, RSV, Strep)	Swab Level	Control Swab Lot # / Exp	Kit Lot # / Exp	Results	Pass? (Y/N)	Initials
1/1/24	Cov	Pos	001234 1/30/25	002345 1/1/26	Pos	Y	JP
1/1/24	Cov	Neg	001235 1/30/25	002345 1/1/26	Neg	Y	JP

Trainee signature:		Date:	
Trainer signature:		Date:	