Training Checklist: **Quidel QuickVue Dipstick Strep A**

INSTRUCTIONS: The Trainee will initial next to each Training Task Description upon reviewing the contents of all checklist items. The Trainer will initial next to each Training Task Description upon training the Trainee. Upon successful completion of this training checklist, the Trainee and Trainer will sign and date the checklist at the end.

Training Task Description	Trainee Initials	Trainer Initials
External Quality Control (QC) a) External QC swabs are tested during with each new lot/shipment. b) The following data must be recorded when QC is run. QC swab type (positive or negative), QC lot number, QC expiration date, QC results, test date, and testing personnel initials. c) There are 2 control bottles that are run, a positive and a negative control. d) The control bottles should be vigorously mixed prior to testing. e) Add 1 drop of control to a prepared tube with Reagent A and B f) Place a clean swab into the tube. Follow patient test procedure.		
Kit and Reagent Handling a) Do not remove Dipsticks from the foil pouch until ready to perform the assay. b) Do not allow the tip of the reagent bottles to com in contact with sample swabs. c) Do not use Reagent B if the solution is green prior to mixing with Reagent A. d) Do not use expired test kits.		
Sample Handling a) Positively identify the patient and label the swab sample. b) Use the swabs supplied in the kit. c) When swabbing throat, be careful not to touch the tongue, sides or top of the mouth. d) Swabs should be tested as soon as possible after collection, up to 72 hours at room temperature.		
Test Procedure a) Know to vertically dispense 3 drops of Reagent A and 3 drops of Reagent B into a clean tube. b) Place patient swab sample into prepared Reagent Tube, roll swab 5x while pressing swab. Keep swab in tube for 1 minute. c) Roll the swab against the Reagent Tube as the swab is removed. d) Remove Dipstick from foil pouch and place into Reagent Tube. e) Read results in 5 minutes. Positive results may appear sooner.		
Results Interpretation a) Any shade of pink-to-red Test Line with the appearance of a blue procedural Control line indicates a positive result. b) A blue procedural Control Line and no pink Test Line is a presumptive negative result. c) If the blue procedural Control Line does not appear, even if any shade of pink-to-red Test Line appears, the result is invalid. A new test should be performed on a new swab and test strip.		
Support a) Know to call Quidel Technical Support when requiring technical assistance, troubleshooting, and error reports – 1 800 874 1517, option 5 or technicalsupport@quidel.com.		

Training Checklist: **Quidel QuickVue Dipstick Strep A**

Trainee to run a <u>positive</u> QC swab and enter results into the Quidel Quickvue QC Log below.		
Trainee to run a <u>negative</u> QC swab and enter results into the Quidel Quickvue QC Log below.		

QC LOG

QUIDEL QuickVue

Date	Test Kit (Cov, Flu, RSV, Strep)	Swab Level	Control Swab Lot # / Exp	Kit Lot # / Exp	Results	Pass? (Y/N)	Initials
1/1/24	Strep	Pos	001234 1/30/25	002345 1/1/26	Pos	Y	JP
1/1/24	Strep	Neg	001235 1/30/25	002345 1/1/26	Neg	Y	JP

Trainee signature:		Date:	
Trainer signature:		Date:	