

RECEIVED 17 OCT 2008

Plate Reader Qualification Certificate

Routine qualification

Certificate ref: PR-08-031

Date: 16 Oct 2008

1 Scope

This certificate records the routine qualification of the plate reader used for the primary potency and kinetic solubility work. It covers the optical checks and the control response checks made on the qualification day.

2 Control response checks

Both control plates were read in the standard layout. The positive and negative control responses were compared against the acceptance record for the instrument.

Table 1 Control response checks

Check	Result	Comment
Positive control response	Within window	Read on both plates
Negative control response	Within window	Low signal, no edge effect
Filter and optics check	Pass	Reference plate read
Plate-to-plate repeatability	Within window	Duplicate plates agreed

3 Qualification status

The plate reader is qualified for use on the primary potency and solubility assays from the date of this certificate. Plates read before qualification on the previous check are covered by the earlier certificate.

- Qualified for primary potency plates
- Qualified for kinetic solubility plates
- Not qualified for the medium-throughput layout until the next check

4 Next review

The next routine qualification is due at the start of the new year. Any repair or lamp change before then requires a new qualification before plates are read.

Name	Role	Signature	Date
Ines Alvarado	Assay Quality Coordinator	<u>I Alvarado</u>	16 Oct 2008