

RECEIVED 17 OCT 2008

Plate Reader Qualification Certificate

Routine qualification

Date: 16 Oct 2008

Instrument identity and qualification references: see signed instrument record

1 Scope

This certificate records routine qualification of the plate reader for primary potency and kinetic solubility work. The signed instrument record identifies the reader, optical settings, plate layouts and qualification run sheets covered by this release.

2 Control response checks

Control-response, optical-check and repeatability records are retained with the signed qualification run sheets. Consult those records for the methods, acceptance limits and observations; this certificate does not replace them.

Table 1 Supporting qualification records

Record	Source
Control response checks	Signed qualification run sheets
Optical checks	Instrument qualification record
Repeatability assessment	Instrument qualification record

3 Qualification status

The plate reader is qualified from the date of this certificate for the methods, optical settings and plate layouts covered by the signed qualification record. This release does not retrospectively qualify earlier runs.

- Primary potency work within the qualified settings
- Kinetic solubility work within the qualified settings
- Confirm that the intended plate layout is covered before use

4 Next review

The next routine qualification is due at the start of the new year. Any repair or lamp change before then requires a new qualification before plates are read.

Name	Role	Signature	Date
------	------	-----------	------

Ines Alvarado	Assay Quality Coordinator	<u>I Alvarado</u>	16 Oct 2008
---------------	------------------------------	-------------------	-------------