

Certificate of calibration**No. 113-3475***Object***1 ring gauge made of steel, \varnothing 40 mm**

Identification: SIP 42923

Order

Calibration of diameter and roundness deviation

*Applicant***SIP, Société Genevoise d'Instruments de Physique**

rue Pré - de - la - Fontaine

CH-1217 Meyrin

Traceability

The reported measurement values are traceable to national standards and thus to internationally supported realizations of the SI-units.

Date of calibration

7 and 10 January 2002



Marking

metas-calibration label

Wabern, 10 January 2002 Sp

For the measurements

Division of Mechanics, Ra


Jürg Spiller
Dr. Bruno Vaucher, Deputy Director

This document may not be published or forwarded other than in full.

Page 1 of 2

Certificate of calibration**No. 113-3475****Extent of calibration**

The diameter and roundness deviation at three heights of the cylinder were measured.

Measurement procedure and conditions

The diameter of the ring gauge was calibrated on a length measurement machine using a laser interferometer and mechanical probing, according to the internal calibration procedure 11370K02.

Measurement probe: ruby sphere, \varnothing 4 mm

Measurement force: extrapolated to zero

Measurement direction: between the reference marks, perpendicular to the cylinder axis

The roundness measurement was carried out on a form measurement machine.

Measurement probe: ruby sphere, \varnothing 4 mm

Measurement force: < 0.05 N

Filter: 2-50 upr, 2RC, phase corrected

The roundness deviation was measured according to ISO 6318. It is defined as the peak to valley deviation from the least squares (LS) circle fitted to the measured profile.

The ambient temperature during the measurements was (20 ± 0.2) °C. The temperature of the ring gauge was 19.90 °C. The diameter results were corrected to the reference temperature of 20 °C assuming a linear coefficient of thermal expansion of $11.5 \cdot 10^{-6}$ K⁻¹.

Measurement results

Identification	Measurement position from mid height	Measured diameter	Roundness
SIP 42923	+7.0 mm	39.999'62 mm	0.06 μ m
	0	39.999'65 mm	0.05 μm
	-7.0 mm	39.999'66 mm	0.06 μ m

Measurement uncertainty


diameter: $U = 0.10$ μ m

roundness: $U = 0.10$ μ m

The reported uncertainty of measurement is stated as the combined standard uncertainty multiplied by a coverage factor $k = 2$. The measured value (y) and the associated expanded uncertainty (U) represent the interval ($y \pm U$) which contains the value of the measured quantity with a probability of approximately 95%. The uncertainty was estimated following the guidelines of the ISO.

The measurement uncertainty contains contributions originating from the measurement standard, from the calibration method, from the environmental conditions and from the object being calibrated.

- Attachment: 4 form measurement protocols



RTH IR300 V04.00 S0

metas, Bern-Wabern

MZ VERT PARALLELISM

Featurename EAM

Measurement no. 04

Par. 0.08 um

P-U 0.03 um

Par Angle 1.0 sec

Trav. Lth. 14.0 mm

Trav. Start 8.4 mm

Trav. End 22.4 mm

Spindle ang. 359.9 deg

Datum EAM 03

Filter 0.80 mm

Profile 100.0 %

Meas. Mode Vert. Up

Meas. date

Meas. time 08:38:34

SIP

Ring gauge, SIP ; D = 40 mm

1217 Meyrin

Ident. No:42923