

**Certificate of calibration****No. 113-3217***Object***1 plug gauge made of steel,  $\varnothing$  50 mm**

Identification: SIP 43046

*Order*

Calibration of diameter and roundness deviation

*Applicant***SIP, Société Genevoise d'Instruments de Physique**

CH-1217 Meyrin

*Traceability*

The reported measurement values are traceable to national standards and thus to the SI-units.

*Date of calibration*

18. April and 3. May 2001


*Marking*

metas-calibration label


CH-3003 Bern-Wabern, 8 May 2001 Sp

For the measurements

Division of Mechanics, Radiation and Legal Metrology



Jürg Spiller



Dr. Bruno Vaucher, Deputy Director



**Certificate of calibration (ctd.)****No. 113-3217****Extent of calibration**

The diameter and roundness deviation at three heights of the cylinder were measured.

**Measurement procedure and conditions**

The diameter of the plug gauge was calibrated on a length measurement machine using a laser interferometer and mechanical probing, according to the internal calibration procedure 11370K02.

Measurement probe: ruby sphere,  $\varnothing$  4 mm

Measurement force: extrapolated to zero

Measurement direction: between the reference marks, parallel to the marking, perpendicular to the cylinder axis

The roundness measurement was carried out on a form measurement machine.

Measurement probe: ruby sphere,  $\varnothing$  4 mm

Measurement force:  $< 0.05$  N

Filter: 2-50 upr, 2RC, phase corrected

The roundness deviation was measured according to ISO 6318. It is defined as the peak to valley deviation from the least squares (LS) circle fitted to the measured profile.

The ambient temperature during the measurements was  $(20 \pm 0.2)$  °C. The temperature of the ring gauge was 19.88 °C. The diameter results were corrected to the reference temperature of 20 °C assuming a linear coefficient of thermal expansion of  $11.6 \cdot 10^{-6}$  K<sup>-1</sup>.

**Measurement results**

Identification	Measurement position from front face of cylinder	Measured diameter	Roundness
SIP 43046	5.0 mm	50.010'25 mm	0.05 $\mu$ m
	12.5 mm	<b>50.010'20 mm</b>	<b>0.09 <math>\mu</math>m</b>
	20.0 mm	50.010'12 mm	0.09 $\mu$ m

**Measurement uncertainty**

diameter:  $U = 0.15$   $\mu$ m

roundness:  $U = 0.10$   $\mu$ m

The reported uncertainty of measurement is stated as the combined standard uncertainty multiplied by a coverage factor  $k = 2$ . The measured value ( $y$ ) and the associated expanded uncertainty ( $U$ ) represent the interval ( $y \pm U$ ) which contains the value of the measured quantity with a probability of approximately 95%. The uncertainty was estimated following the guidelines of the ISO.


The measurement uncertainty contains contributions originating from the measurement standard, from the calibration method, from the environmental conditions and from the object being calibrated.

- Attachment: 4 form measurement protocols

This document may not be published or forwarded other than in full.

Page 2 of 2





RTH TR300 V04.00 S0

metas, Bern-Wabern

MZ VERT PARALLELISM

Featurename EAMD

Measurement no. 04

Par. 0.26 um

P-U 0.05 um

Par Angle -3.2 sec

Trav. Lth. 15.0 mm

Trav. Start 85.5 mm

Trav. End 100.5 mm

Spindle ang. 0.0 deg

Datum EAMD 03

Filter 0.80 mm

Profile 100.0 %

Meas. Mode Vert. Up

Meas. date

Meas. time 20:20:50

SIP

Pflug gauge , SIP , D = 50 mm

1217 Meyrin

Ident. No:43046