

Alconox, Inc.

The Leader in Critical-Cleaning

30 Glenn St., Suite 309, White Plains, NY 10603 USA
Tel.914.948.4040 ● Fax.914.948.4088
www.alconox.com ● cleaning@alconox.com

Factory: 61 Cornelison Ave. Jersey City, NJ 07304 USA

Date COA Created: 9/26/2017

CERTIFICATE OF ANALYSIS CITRAJET

Manufactured by: Alconox, Inc.

Sub lot No: H1J7 H1J7F, H1J7G, H1J7J, H1J7K, H1J7L, H1J7M

Sub lot numbers of Master lot No: H1H6

Manuf. Cat. no.: 2001, 2001-1, 2005, 2015, 2030, 2055

Date Of Manufacture: 9/22/2017

Expires end of: 09/19

Storage Conditions: Closed upright and in a cool dry place (should be 15 - 30 deg C or 60 - 90 deg F).

ANALYSIS	RESULT	ACCEPTED RANGE
Appearance:	OK	Clear, olive colored liquid
pH: (1% solution in DI water)	3.1	2.1 - 3.4
Surface Tension:(1% solution Dyne/cm)	33.4	29.0 - 35.0
Specific Gravity:	1.126	1.09 - 1.15
Foam 1% Solution (110 Deg F)	NIL	NIL

Malcolm McLaughlin, QA

Alconox, Inc.

mm

Contains no class 1, class 2, class 3, or class 4 residual solvent ingredients as defined by ICH Q3C guidelines.

COA NUMBER: 7351