



Alconox, Inc.

The Leader in Critical-Cleaning

30 Glenn St., Suite 309, White Plains, NY 10603 USA

Tel.914.948.4040 • Fax.914.948.4088

www.alconox.com • cleaning@alconox.com

Factory: 61 Cornelison Ave. Jersey City, NJ 07304 USA

Date COA Created: 6/20/2018

CERTIFICATE OF ANALYSIS

CITRAJET

Manufactured by: Alconox, Inc.

Sub lot No: H1F8 H1F8F, H1F8G, H1F8J, H1F8K, H1F8L, H1F8M

Sub lot numbers of Master lot No: H1L7

Manuf. Cat. no.: 2001, 2001-1, 2005, 2015, 2030, 2055

Date Of Manufacture: 6/19/2018

Expires end of: 06/20

Storage Conditions: Closed upright and in a cool dry place (should be 15 - 30 deg C or 60 - 90 deg F).

ANALYSIS

RESULT

ACCEPTED RANGE

Appearance:	OK	Clear, olive colored liquid
pH: (1% solution in DI water)	3.1	2.1 - 3.4
Surface Tension:(1% solution Dyne/cm)	33.5	29.0 - 35.0
Specific Gravity:	1.132	1.09 - 1.15
Foam 1% Solution (110 Deg F)	NIL	NIL

Malcolm McLaughlin, QA

Alconox, Inc.

Contains no class 1, class 2, class 3, or class 4 residual solvent ingredients as defined by ICH Q3C guidelines.

COA NUMBER: 7547