



Alconox, Inc.

Critical Cleaning Experts

30 Glenn St., Suite 309, White Plains, NY 10603 USA

Tel.914.948.4040 • Fax.914.948.4088

www.alconox.com • cleaning@alconox.com

Factory: 61 Cornelison Ave. Jersey City, NJ 07304 USA

Date COA Created: 8/4/2020

## **CERTIFICATE OF ANALYSIS**

### **CITRAJET**

**Manufactured by: Alconox, Inc.**

Sub lot No: H1G0 H1G0F, H1G0G, H1G0J, H1G0K, H1G0L, H1G0M

Sub lot numbers of Master lot No: H1F9

Manuf. Cat. no.: 2001, 2001-1, 2005, 2015, 2030, 2055

Date Of Manufacture: 7/30/2020

Expires end of: 07/22

Storage Conditions: Closed upright and in a cool dry place (should be 15 - 30 deg C or 60 - 90 deg F).

#### **ANALYSIS**

#### **RESULT**

#### **ACCEPTED RANGE**

Appearance:	OK	Clear, olive colored liquid
pH: (1% solution in DI water)	3.1	2.1 - 3.4
Surface Tension:(1% solution Dyne/cm)	33.3	29.0 - 35.0
Specific Gravity:	1.13	1.09 - 1.15
Foam 1% Solution (110 Deg F)	NIL	NIL

**Laura Falla, QA**

**Alconox, Inc.**

*Contains no class 1, class 2, class 3, or class 4 residual solvent ingredients as defined by ICH Q3C guidelines.*