

Alconox, Inc.

Critical Cleaning Experts

30 Glenn St., Suite 309, White Plains, NY 10603 USA Tel.914.948.4040 - Fax.914.948.4088 www.alconox.com - cleaning@alconox.com

Factory: 61 Cornelison Ave. Jersey City, NJ 07304 USA

Date COA Created: 3/1/2019

CERTIFICATE OF ANALYSIS LIQUINOX

Manufactured by: Alconox, Inc.

Sub lot No: L3B9 L3B9F, L3B9H, L3B9J, L3B9K, L3B9L, L3B9M

Sub lot numbers of Master lot No: L1H7

Manuf. Cat. no.: 1201, 1201-1, 1205, 1215, 1230, 1232, 1232-1, 1255

Date Of Manufacture: 2/27/2019

Expires end of: 02/21

Storage Conditions: Closed upright and in a cool dry place (should be 15 - 30 deg C or 60 - 90 deg F).

| ANALYSIS | <u>RESULT</u> | ACCEPTED RANGE |
|---------------------------------------|---------------|---------------------|
| Appearance: | OK | Clear golden liquid |
| pH: (as is) | 8.5 | 8.0 - 9.5 |
| Surface Tension:(1% solution Dyne/cm) | 30.3 | 27.0 - 31.0 |
| Specific Gravity: | 1.08 | 1.06 - 1.09 |

Laura Falla, QA

Janua Falla

Alconox, Inc.

Contains no class 1, class 2, class 3, or class 4 residual solvent ingredients as defined by ICH Q3C guidelines.

COA NUMBER: 7730