

CERTIFICATE OF ANALYSIS
CITRANOX
FOR BECKMAN COULTER (BCI)

Date Of Manufacture: 07/25/2016

Manufactured by: Alconox, Inc.

Manufacturer Lot number: C9G6

BCI Lot number: 630000


Manuf. Cat. No.: 1801B

BCI Part No: 81912

Shelf Life Expiration Date: 06/30/2018 To meet BC requirements


Storage Conditions: Closed upright and in a cool dry place (15 - 30 deg C or 60 - 90 deg F).

<u>Value</u>	<u>-</u>	<u>Result</u>	<u>Acceptance Range</u>
Appearance:		OK	Clear pale olive colored liquid
pH: (1% solution in DI water)		2.8	2.0-3.0
Surface Tension: (1% solution Dyne/cm)		31.2	29.0 – 33.0
Specific Gravity		1.112	1.090 – 1.140
Number of Bar Code Labels Used for Labels		720	Not Applicable
Number of Bar Code Labels Used for Records		3	Not Applicable
Number of Bar Code Labels Discarded		37	Not Applicable



Malcolm McLaughlin, QA

REF 81912

 **2018-06-30**

LOT 630000



+H628819121A



+\$\$2063018630000AT



(01)15099590223571

(17)180630

(10)630000

Alconox, Inc.

The Leader in Critical Cleaning

30 Glenn St. White Plains NY 10603 USA

Tel 914.948.4040 • Fax 914-948-4088

www.alconox.com • cleaning@alconox.com