

IoT- Environmental Monitoring

NAME : Vibisha. W

ID:aut962921104712

EMAIL:vv3956601@gmail.com




TABLE OF CONTENTS



Innovation.....	3
User Understanding.....	6
Conclusion.....	7

INNOVATION:

Temperature and Humidity:


Relative Humidity is a characteristic of air, and indicates a measure of the amount of moisture in the air. When the relative humidity is high, the air feels clammy or muggy. We may sweat more and have a hard time staying cool. When the relative humidity is low, our lips get chapped, skin dries and cracks, and things contacting our skin feel a little rougher. Other living (or once-living) things like plants, insects, mold and wood also react to changes in relative humidity.

Relative humidity is a ratio or the “relative” amount of moisture in the air, as compared to the maximum amount of moisture the air can hold **at that temperature**. For example, air with a relative humidity of 50% contains one half the amount of moisture the air can hold at that temperature. Warmer air can hold more moisture than colder air, so when air is warmed, its relative humidity decreases. Air at 70 degrees and 50% RH will have a RH of only 25% when heated to 90 degrees. Conversely, the 70 degree, 50% RH air will have a 100% relative humidity when cooled to 50 degrees. The following chart shows the changes in relative humidity as 90 degree/ 50% RH air is cooled. At about 76 degrees, the RH becomes 80%, allowing mold to grow. At about 69 degrees, condensation starts to form.

Controlling relative humidity in a building is important because of its effects on both the structure and living organisms within the building. People are most comfortable in a relative humidity of 30-60% while mold grows best at relative humidity over 80%. Dust mites like a relative humidity over 60%, and can be controlled by maintaining a RH below 50%. Wood expands at higher RH, and shrinks at low RH. Situations where hardwood floors cup, warp or gap are usually caused by RH problems. An ideal situation would be a stable indoor relative humidity between 30 and 50% all year long.

Relative Humidity and Temperature Relationship

Cooling 90 F and 50% RH Air



How to Determine relative humidity

Several methods are available for determining relative humidity (RH). The easiest method is to purchase a digital hygrometer, or thermo-hygrometer. These devices can cost less than \$50, and often provide temperature as well as RH readouts. Though these inexpensive devices are not the most accurate, they can show trends and indications of potential problems. More accurate devices can cost \$100 and up.

These inexpensive hygrometers can be calibrated (more accurately “interpreted”) for more accurate readings. To calibrate, fill a small soup bowl 2/3’s full with tap water. Add table salt and stir until no more salt dissolves, and some salt crystals remain in the bottom of the bowl. Add another teaspoon of salt just to be sure. Place the bowl of salty water and the hygrometer in a large plastic bag (Make sure the hygrometer does not get wet), seal the bag shut, and read the RH after about 6 hours. The reading should be about 75% RH. If not, add or subtract the difference as a calibration factor. For example, if your hygrometer reads 70%, you should add 5% to all readings. If the reading was 85%, you should subtract 10% from all readings.

A second method for determining relative humidity is by using a simple thermometer, a glass of ice water and a psychometric chart.

Step 1. Measure the temperature of the air with the thermometer.

Step 2. Fill a metal or thin glass drinking glass 2/3’s full with tap water. Add a few ice cubes and stir with the thermometer.

Step 3. Watch the outside of the glass as you stir. As soon as you see condensation (a haze) form on the outside of the glass, read the thermometer.

Step 4. Find the two temperatures on a psych chart to determine the relative humidity from your readings.

User Understanding:

HUMAN TEMPERATURE											
Relative Humidity	Air Temperature (F°)										
	70°	75°	80°	85°	90°	95°	100°	105°	110°	115°	
	Apparent temperature										
10%	65°	70°	75°	80°	85°	90°	95°	100°	105°	111°	
20%	66°	72°	77°	82°	87°	93°	99°	105°	112°	120°	
30%	67°	73°	78°	84°	90°	96°	104°	113°	123°	135°	
40%	68°	74°	79°	86°	93°	101°	110°	123°	137°		
50%	69°	75°	81°	88°	96°	107°	120°	135°	150°		
60%	70°	76°	82°	90°	100°	114°	132°	149°			
70%	70°	77°	85°	93°	106°	124°	144°				
80%	71°	78°	86°	97°	113°	136°	157°				
90%	71°	79°	88°	102°	122°	150°	170°				
100%	72°	80°	91°	108°	133°	166°					

This chart tells us everything we need to know about how humidity affects our perception of temperature.

Example: Let's say that we have a 75°F temperature in a house. That means that if we check the thermometer, it will always read '75°F'. However, we won't always **feel** the indoor air at 75°F. Humidity plays a big role in how we perceive air temperature. Namely, we will perceive 75°F temperature as:

- 70°F at 10% humidity levels.

- 72°F at 20% humidity levels.
- 73°F at 30% humidity levels.
- 74°F at 40% humidity levels.
- 75°F at 50% humidity levels.
- 76°F at 60% humidity levels.
- 77°F at 70% humidity levels.
- 78°F at 80% humidity levels.
- 79°F at 90% humidity levels.
- 80°F at 100% humidity levels.

As we can see from the temperature and humidity relationship chart, 75°F can feel like 70°F at very low humidity levels (10% moisture levels) or 80°F at very high humidity levels (100% moisture levels). This is a full 10°F range; meaning that humidity levels have a substantial effect on the temperature levels we feel.

Let's first look at why our perception of how hot/cold air depends on humidity. After that, we will look in-depth at how humidity affects the temperature based on the Temperature And Humidity Relationship Chart.

Conclusion:

The presence of water vapour in the atmosphere is known as humidity. Hence more water evaporates in a given location, the more water vapour raises into the air, raising the humidity. Because temperature affects water to evaporate faster, hot regions are more humid than cool places.

