



ND150D



SL.NO.	DESCRIPTION
01.	VALVE BODY.
02.	ROUGHING CONNECTION.
03.	JET ASSEMBLY
04.	BACKING CONNECTION
05.	HEATER
06.	FLUID FILLER AND DIPSTICK ASSY.
07.	FLUID DRAIN
08.	CRYO COOLED LNT



NOTE: 1) Hydrostatic test to be done at a pressure of 4 kg/cm² for cooling coils
2) No. of cooling coils are 16

		HIND HIGH VACUUM CO.(P) LTD. SITE NO:17,PEENYA INDL AREA,PHASE-1 BANGALORE 560 058																				
General tolerances where not specified		02-24-000-000 1.12-02-14-000-000																				
Linear Dimensions																						
Over	Upto	Tol.																				
-	10	±0.1																				
10	30	±0.2																				
30	100	±0.3	700 & 305 Wht 710 & 324 -REV 1													9/11	1/2					
100	300	±0.5	ALTERATION												INDEX	AREA	DATE	NAME	QTY.	Assy.No.		
300	1000	±0.8												Scale	Drawn	12.04.02		Projection				
1000	2000	±1.2												NTS	Check	12.04.02	1/2					
Angular Dimensions		ITEM	MATERIAL	WEIGHT	PATTERN NO.	REMARKS								Appvd								
Length of Shorter arm		DESIGNATION																DRAWING NO.		SHEET NO.		
Over	Upto	Tol.	DIFFUSION PUMP OD150D																02-14-000-000		01/01	
-	10	±1°	DIFFSTACK TYPE																REV-01			
10	50	±30'																				
50	100	±20'																				
100	-	±10'																				