


ETDS QA Overview - L1 | Pads RT°C | Couch 0deg





Sphere detection comparison - vectors as (x, y, z) in mm

Defl.	SD Reference Coordinates	SD Movement Vector	Phantom Movement Vector	Movement Error (SD - Phantom)	Surface Tracking Vector
1	(-0.30, 0.30, 0.20)	(-0.20, -18.10, -0.70)	(0.00, -17.72, -0.69)	(-0.20, -0.38, -0.01)	(-0.10, -17.90, -0.60)
2	(-0.20, 0.20, 0.10)	(-0.30, -37.50, -3.10)	(0.00, -37.31, -3.49)	(-0.30, -0.19, 0.39)	(-0.40, -37.30, -3.30)

Sphere detection - absolute deviations

ETDS QA Overview - L1 | Pads RT°C | Couch 0deg

Lateral (X)



Defl.	Deviation	Lateral (X) [mm]	Longitudinal (Y) [mm]	Vertical (Z) [mm]
D1	SD - Phantom	0.20	0.38	0.01
D1	SD - Surface Tracking	0.10	0.20	0.10
D2	SD - Phantom	0.30	0.19	0.39
D2	SD - Surface Tracking	0.10	0.20	0.20

Absolute deviations of the X-ray based sphere detection from the phantom axis setpoint and from the ETD surface tracking, evaluated at the deflected position (end of measurement window). Tolerance rings: accept 1 mm, watch 2 mm.