


ETDS QA Overview - L3 | Pads 32°C | Couch 0deg





Sphere detection comparison - vectors as (x, y, z) in mm

Defl.	SD Reference Coordinates	SD Movement Vector	Phantom Movement Vector	Movement Error (SD - Phantom)	Surface Tracking Vector
1	(0.50, 0.60, 0.20)	(0.00, -17.80, -0.70)	(0.00, -17.72, -0.69)	(0.00, -0.08, -0.01)	(0.10, -17.50, -0.70)
2	(0.50, 0.70, 0.20)	(0.10, -37.30, -3.40)	(0.00, -37.31, -3.49)	(0.10, 0.01, 0.09)	(0.19, -37.01, -3.59)

Sphere detection - absolute deviations

ETDS QA Overview - L3 | Pads 32°C | Couch 0deg

Lateral (X)



Defl.	Deviation	Lateral (X) [mm]	Longitudinal (Y) [mm]	Vertical (Z) [mm]
D1	SD - Phantom	0.00	0.08	0.01
D1	SD - Surface Tracking	0.10	0.30	0.00
D2	SD - Phantom	0.10	0.01	0.09
D2	SD - Surface Tracking	0.09	0.29	0.19

Absolute deviations of the X-ray based sphere detection from the phantom axis setpoint and from the ETD surface tracking, evaluated at the deflected position (end of measurement window). Tolerance rings: accept 1 mm, watch 2 mm.