


ETDS QA Overview - L3 | Pads RT°C | Couch 0deg




Sphere detection comparison - vectors as (x, y, z) in mm

Defl.	SD Reference Coordinates	SD Movement Vector	Phantom Movement Vector	Movement Error (SD - Phantom)	Surface Tracking Vector
1	(0.50, 0.80, 0.20)	(0.10, -18.00, -0.70)	(0.00, -17.72, -0.69)	(0.10, -0.28, -0.01)	(0.10, -17.70, -0.70)
2	(0.50, 0.70, 0.20)	(0.10, -37.30, -3.40)	(0.00, -37.31, -3.49)	(0.10, 0.01, 0.09)	(0.20, -36.98, -3.60)

Sphere detection - absolute deviations

ETDS QA Overview - L3 | Pads RT°C | Couch 0deg

Lateral (X)



	D1: SD - Phantom		D2: SD - Phantom
	D1: SD - Surface Tracking		D2: SD - Surface Tracking

Defl.	Deviation	Lateral (X) [mm]	Longitudinal (Y) [mm]	Vertical (Z) [mm]
D1	SD - Phantom	0.10	0.28	0.01
D1	SD - Surface Tracking	0.00	0.30	0.00
D2	SD - Phantom	0.10	0.01	0.09
D2	SD - Surface Tracking	0.10	0.32	0.20

Absolute deviations of the X-ray based sphere detection from the phantom axis setpoint and from the ETD surface tracking, evaluated at the deflected position (end of measurement window). Tolerance rings: accept 1 mm, watch 2 mm.