


ETDS QA Overview - L4 | Pads RT°C | Couch 0deg





Sphere detection comparison - vectors as (x, y, z) in mm

Defl.	SD Reference Coordinates	SD Movement Vector	Phantom Movement Vector	Movement Error (SD - Phantom)	Surface Tracking Vector
1	(-1.10, -0.60, -0.30)	(0.20, -17.80, -0.70)	(0.00, -17.72, -0.69)	(0.20, -0.08, -0.01)	(0.20, -17.58, -0.50)
2	(-1.10, -0.50, -0.30)	(0.30, -37.20, -3.30)	(0.00, -37.31, -3.49)	(0.30, 0.11, 0.19)	(0.30, -36.90, -3.16)

Sphere detection - absolute deviations

ETDS QA Overview - L4 | Pads RT°C | Couch 0deg

Lateral (X)



Defl.	Deviation	Lateral (X) [mm]	Longitudinal (Y) [mm]	Vertical (Z) [mm]
D1	SD - Phantom	0.20	0.08	0.01
D1	SD - Surface Tracking	0.00	0.22	0.20
D2	SD - Phantom	0.30	0.11	0.19
D2	SD - Surface Tracking	0.00	0.30	0.14

Absolute deviations of the X-ray based sphere detection from the phantom axis setpoint and from the ETD surface tracking, evaluated at the deflected position (end of measurement window). Tolerance rings: accept 1 mm, watch 2 mm.