

3004FinalProject

Team 18: TA Deen

Authors:

Julien Rougerie 101067032

Rinith Anand 100906754

Aaron Saville 101117379

Andrew Hall 101156505

Files Included:

QT: mainwindow.ui, CESDEVICE.pro, AlphaTest.pro, AlphaTest.pro.user, guiAssets.qrc

Headers: mainwindow.h, battery.h, countdownclock.h, record.h

CPP: main.cpp, mainwindow.cpp, countdownclock.cpp, battery.cpp, record.cpp.

Special Files: 12 sequence diagrams, 1 UML diagram, this readme, testcases pdf, usecases pdf.

auto_off_SequenceDiagram.pdf, batteryDrain_SequenceDiagram.pdf,

lockSequence_SequenceDiagram.pdf, OverLoadShutoff_SequenceDiagram.pdf,

record_therapy_SequenceDiagram.pdf, SelectFrequencySequenceDiagram.pdf,

select_wavelength_SequenceDiagram.pdf, set_current_SequenceDiagram.pdf,

StartSession_SequenceDiagram.pdf, TherapyTimer_SequenceDiagram.pdf,

turn_Device_off_SequenceDiagram.pdf, turn_device_on_SequenceDiagram.pdf,

UML_CRA_F.pdf

This project was a joint collaboration, almost all the files were touched by everyone at one point in time as implementations were created, changed and refined.

Specific assignments were chosen by members via website trello which is decently accurate:

<https://trello.com/invite/b/85Esv1RZ/0635d05e16dac009ed82425acf6f98a7/comp-3004>

You can check the github changelog for more info as well.

The files found in the main folder represents all of the Qt and c++ code.

The Documentation Folder holds all documentation raws and the Folders with assets in the name hold all assets for the CRA-Device.

Documentation pdfs holds the pdf files of the required documentation.

Tested Scenarios:

See the included test cases file.

Notes:

You will notice there are two power modes with a checkbox labeled "realistic power draw", this is designed as to allow every therapy to be completed at least once without the device running out of power. If you uncheck this box the device will not function as intended as the power draw will be unrealistically high. This feature was included just for the demo so the TA could watch the power decrease faster.

You will also notice that our device drains power even when there is not a therapy active, this was a design decision to be more realistic. Undergoing a therapy will however increase the amount of power being drained at any given time as required.

We opted for the therapy history to store the last active current used in the therapy.

The autoshut off is set to 30 seconds instead of 30mins so as to not waste your time.

A power overload will put the device in an irrecoverable state and you will need to relaunch the program if you want to continue after an overload. This was intended as the device is being simulated to have shorted itself internally.