


La Jolla, CA; AMF1 (main site for EPCAPE on Scripps Pier)
Atmospheric Radiation Measurement User Facility



ARM Data Quality Report (DQR) Table

Datastream	DQR	Quality	Subject	Start Date	End Date
epcaosacsmM1.b1	D230923.1	incorrect	Mentors On-Site for Mid-Campaign Calibration	2023-09-18 00:00:00	2023-09-22 23:59:59
epcaosacsmM1.b1	D231129.2	incorrect	Incorrect Data	2023-11-20 04:00:00	2023-11-22 16:59:18
epcaosacsmM1.b1	D240215.2	incorrect	Incorrect Data	2024-02-09 01:00:00	2024-02-09 03:00:00
epcaosacsmM1.b1	D231129.1	missing	Missing Data	2023-11-22 16:59:18	2023-11-24 16:39:20
epcaosaeth2spotM1.b1	D230817.4	incorrect	Instrument Failure	2023-08-04 20:40:00	2023-08-27 18:00:00
epcaosaeth2spotM1.b1	D230923.1	incorrect	Mentors On-Site for Mid-Campaign Calibration	2023-09-18 00:00:00	2023-09-22 23:59:59
epcaosaeth2spotM1.b1	D231114.4	missing	Data Not Available	2023-10-15 08:59:01	2023-10-20 21:57:59
epcaosmetM1.a1	D230923.1	incorrect	Mentors On-Site for Mid-Campaign Calibration	2023-09-18 00:00:00	2023-09-22 23:59:59
epcaosmetM1.a1	D231114.9	missing	Data Not Available	2023-10-15 09:00:00	2023-10-20 19:31:49
epcaosmetM1.a1	D240208.2	suspect	Precipitation Data Suspect	2024-02-01 13:46:00	2024-02-01 21:46:59
epcaosapsM1.b1	D230302.2	incorrect	Revision Required	2023-02-15 00:00:00	2023-03-09 23:59:59
epcaosapsM1.b1	D230510.2	incorrect	Instrument Undergoing On-Site Check	2023-05-10 16:00:00	2023-05-10 17:59:59
epcaosapsM1.b1	D230514.1	incorrect	Instrument Undergoing On-Site Check	2023-05-13 15:00:00	2023-05-14 18:30:00
epcaosapsM1.b1	D230602.2	incorrect	Data Incorrect	2023-06-01 20:00:00	2023-06-01 21:00:00
epcaosapsM1.b1	D230913.1	incorrect	Instrument Flow disrupted due to Power Outage	2023-09-12 23:00:00	2023-09-13 00:15:00
epcaosapsM1.b1	D230923.1	incorrect	Mentors On-Site for Mid-Campaign Calibration	2023-09-18 00:00:00	2023-09-22 23:59:59
epcaosapsM1.b1	D230625.1	suspect	Suspicious Data	2023-05-01 18:00:00	2023-05-12 07:00:00
epcaosapsM1.b1	D230702.1	suspect	Suspicious Data	2023-04-01 00:00:00	2023-04-20 00:00:00
epcaosapsM1.b1	D230702.3	suspect	Suspicious Data	2023-03-01 00:00:00	2023-03-31 23:59:59
epcaosapsM1.b1	D230929.4	missing	Instrument Down	2023-09-18 21:07:42	2023-10-23 22:23:35
epcaosccn2colaM1.b1	D230825.2	suspect	Clock Synchronization Issue	2023-05-27 01:04:19	2023-05-30 18:22:06
epcaosccn2colaM1.b1	D230923.1	incorrect	Mentors On-Site for Mid-Campaign Calibration	2023-09-18 00:00:00	2023-09-22 23:59:59
epcaosccn2colaM1.b1	D231019.1	incorrect	Instrument Flooded	2023-09-29 09:00:00	2023-11-30 00:00:00
epcaosccn2colaM1.b1	D240103.2	incorrect	Incorrect Flow Ratio for CCN Col A	2023-12-28 15:00:00	2023-12-28 20:00:00
epcaosccn2colaM1.b1	D240130.1	incorrect	Leak in the CCN Sample Line	2023-02-15 00:00:00	2024-02-15 00:00:00
epcaosccn2colaM1.b1	D240205.3	incorrect	Incorrect Data	2024-01-31 20:00:00	2024-01-31 23:00:00
epcaosccn2colaM1.b1	D231215.1	missing	Data Not Available	2023-10-13 22:36:58	2023-11-01 18:20:03
epcaosccn2colaM1.b1	D231215.4	missing	Instrument Flooded	2023-10-07 01:08:09	2023-11-21 00:30:53
epcaoscoM1.b1	D230923.1	incorrect	Mentors On-Site for Mid-Campaign Calibration	2023-09-18 00:00:00	2023-09-22 23:59:59
epcaoscoM1.b1	D231114.5	missing	Data Not Available	2023-10-15 09:00:00	2023-10-20 23:09:43

ARM Data Quality Report (DQR) Table

Datastream	DQR	Quality	Subject	Start Date	End Date
epcaospcpu1mM1.b1	D230322.7	suspect	Vacuum Supply to the Instrument is Sub-Optimal	2023-03-22 00:00:00	2023-03-22 16:00:00
epcaospcpu1mM1.b1	D230913.1	incorrect	Instrument Flow disrupted due to Power Outage	2023-09-12 23:00:00	2023-09-13 00:15:00
epcaospcpu1mM1.b1	D230923.1	incorrect	Mentors On-Site for Mid-Campaign Calibration	2023-09-18 00:00:00	2023-09-22 23:59:59
epcaospcpu1mM1.b1	D240102.1	incorrect	Instrument Undergoing Service	2024-01-02 14:20:00	2024-01-02 15:00:00
epcaospcpu1mM1.b1	D240205.3	incorrect	Incorrect Data	2024-01-31 20:00:00	2024-01-31 23:00:00
epcaospcpu1mM1.b1	D231114.6	missing	Data Not Available	2023-10-15 08:59:31	2023-10-20 21:00:29
epcaoshtdmaM1.a1	D230322.4	suspect	Sample Flow Fluctuating	2023-02-15 00:00:00	2023-03-04 07:00:00
epcaoshtdmaM1.a1	D230322.5	incorrect	Water in the Instrument	2023-03-04 07:00:00	2023-03-20 20:00:00
epcaoshtdmaM1.a1	D230425.9	incorrect	Instrument not Sampling Ambient Air	2023-04-23 16:00:00	2023-04-24 13:00:00
epcaoshtdmaM1.a1	D230923.1	incorrect	Mentors On-Site for Mid-Campaign Calibration	2023-09-18 00:00:00	2023-09-22 23:59:59
epcaoshtdmaM1.a1	D231218.5	incorrect	Flows not Nominal	2023-12-12 09:00:00	2023-12-16 00:00:00
epcaoshtdmaM1.a1	D230412.6	missing	Instrument Offline for Troubleshooting	2023-03-24 09:00:00	2023-04-11 17:30:00
epcaosnephwet1mM1.b1	D230519.1	suspect	RH Cycling not Nominal	2023-03-17 16:00:00	2023-09-23 00:00:00
epcaosnephwet1mM1.b1	D230923.1	incorrect	Mentors On-Site for Mid-Campaign Calibration	2023-09-18 00:00:00	2023-09-22 23:59:59
epcaosnephwet1mM1.b1	D231114.10	missing	Data Not Available	2023-10-15 08:59:31	2023-10-24 18:00:29
epcaosnephwet1mM1.b1	D231206.3	missing	Data Missing due to Computer Failure	2023-11-20 09:59:58	2023-11-28 15:40:39
epcaoso3M1.b1	D230923.1	incorrect	Mentors On-Site for Mid-Campaign Calibration	2023-09-18 00:00:00	2023-09-22 23:59:59
epcaoso3M1.b1	D231114.14	missing	Data Not Available	2023-10-15 09:00:00	2023-10-20 19:37:39
epcaospsap3w1mM1.b1	D230918.3	incorrect	Incorrect Data	2023-09-11 16:36:20	2023-09-13 01:41:21
epcaospsap3w1mM1.b1	D230923.1	incorrect	Mentors On-Site for Mid-Campaign Calibration	2023-09-18 00:00:00	2023-09-22 23:59:59
epcaospsap3w1mM1.b1	D231114.16	missing	Data Not Available	2023-10-15 09:59:31	2023-10-20 19:00:29
epcaossmpsM1.b1	D230228.1	suspect	Questionable Data	2023-02-17 00:00:00	2023-02-17 06:45:01
epcaossmpsM1.b1	D230228.3	suspect	Instrument Sample Collection Disrupted	2023-02-22 00:00:00	2023-02-25 23:59:59
epcaossmpsM1.b1	D230322.7	suspect	Vacuum Supply to the Instrument is Sub-Optimal	2023-03-22 00:00:00	2023-03-22 16:00:00
epcaossmpsM1.b1	D230702.4	suspect	Instrument Error	2023-03-22 03:00:00	2023-03-22 16:00:00
epcaossmpsM1.b1	D230302.4	incorrect	Instrument Under Review - With Period of No Data	2023-03-01 15:00:00	2023-03-01 23:30:00
epcaossmpsM1.b1	D230510.2	incorrect	Instrument Undergoing On-Site Check	2023-05-10 16:00:00	2023-05-10 17:59:59
epcaossmpsM1.b1	D230514.1	incorrect	Instrument Undergoing On-Site Check	2023-05-13 15:00:00	2023-05-14 18:30:00
epcaossmpsM1.b1	D230702.5	incorrect	Instrument Undergoing Weekly Check	2023-03-20 18:00:00	2023-03-27 20:00:00
epcaossmpsM1.b1	D230913.1	incorrect	Instrument Flow disrupted due to Power Outage	2023-09-12 23:00:00	2023-09-13 00:15:00

ARM Data Quality Report (DQR) Table

Datastream	DQR	Quality	Subject	Start Date	End Date
epcaossmmpsM1.b1	D230916.1	incorrect	Incorrect Data	2023-09-12 23:00:00	2023-09-13 03:00:00
epcaossmmpsM1.b1	D230923.1	incorrect	Mentors On-Site for Mid-Campaign Calibration	2023-09-18 00:00:00	2023-09-22 23:59:59
epcaossmmpsM1.b1	D231009.11	incorrect	Incorrect Data	2023-09-11 16:36:00	2023-09-11 18:24:00
epcaossmmpsM1.b1	D231217.1	incorrect	Instrument Undergoing Weekly Check	2023-10-30 20:00:00	2023-12-04 19:00:00
epcaossmmpsM1.b1	D240127.1	incorrect	Instrument Undergoing Weekly Check	2024-01-24 16:00:00	2024-01-24 18:00:00
epcaossmmpsM1.b1	D240205.3	incorrect	Incorrect Data	2024-01-31 20:00:00	2024-01-31 23:00:00
epcaossmmpsM1.b1	D230313.3	missing	Data Not Available	2023-03-10 18:00:00	2023-03-10 22:00:00
epcaossmmpsM1.b1	D231114.18	missing	Data Not Available	2023-10-15 08:55:01	2023-10-30 15:24:59
epcaosso2M1.b1	D230923.1	incorrect	Mentors On-Site for Mid-Campaign Calibration	2023-09-18 00:00:00	2023-09-22 23:59:59
epcaosso2M1.b1	D231114.20	missing	Data Not Available	2023-10-15 10:00:00	2023-10-20 19:48:14
epcaossp2bc60sM1.b1	D230613.8	missing	No Data	2023-05-23 17:00:00	2023-06-03 10:23:00
epcaossp2bc60sM1.b1	D230721.10	missing	Missing Data	2023-07-08 16:36:00	2023-07-12 15:50:00
epcaossp2bc60sM1.b1	D230923.1	incorrect	Mentors On-Site for Mid-Campaign Calibration	2023-09-18 00:00:00	2023-09-22 23:59:59
epcaosuhsasM1.b1	D230718.1	suspect	Artifacts in the UHSAS data above ~550 nm	2023-02-15 00:00:00	2023-06-10 23:59:54
epcaosuhsasM1.b1	D230923.1	incorrect	Mentors On-Site for Mid-Campaign Calibration	2023-09-18 00:00:00	2023-09-22 23:59:59
epcaosuhsasM1.b1	D231013.2	incorrect	Incorrect Data	2023-09-11 16:37:50	2023-09-11 18:22:19
epcaosuhsasM1.b1	D240205.3	incorrect	Incorrect Data	2024-01-31 20:00:00	2024-01-31 23:00:00

ARM Data Object Identifier (DOI) Table

Instrument	DOI
ACSM	Zawadowicz, M., Watson, T., Hayes, C., Allain, M., Salwen, C., & Behrens, B. Aerosol Chemical Speciation Monitor (AOSACSM), 2023-02-15 to 2024-02-14, ARM Mobile Facility (EPC), La Jolla, CA; AMF1 (main site for EPCAPE on Scripps Pier) (M1). Atmospheric Radiation Measurement (ARM) User Facility. https://doi.org/10.5439/1558768
AETH	Chand, D., Ermold, B., Andrews, E., Sedlacek, A., Hayes, C., & Salwen, C. Aethalometer (AOSAETH2SPOT1M), 2023-02-15 to 2024-02-14, ARM Mobile Facility (EPC), La Jolla, CA; AMF1 (main site for EPCAPE on Scripps Pier) (M1). Atmospheric Radiation Measurement (ARM) User Facility. https://doi.org/10.5439/1991930
AOSMET	Kyroutac, J., & Tuftedal, M. Meteorological Measurements associated with the Aerosol Observing System (AOSMET), 2023-02-15 to 2024-02-14, ARM Mobile Facility (EPC), La Jolla, CA; AMF1 (main site for EPCAPE on Scripps Pier) (M1). Atmospheric Radiation Measurement (ARM) User Facility. https://doi.org/10.5439/1984920
APS	Singh, A., Oliveira, D., Koontz, A., Sedlacek, A., & Kuang, C. Aerodynamic Particle Sizer (AOSAPS), 2023-02-15 to 2024-02-14, ARM Mobile Facility (EPC), La Jolla, CA; AMF1 (main site for EPCAPE on Scripps Pier) (M1). Atmospheric Radiation Measurement (ARM) User Facility. https://doi.org/10.5439/1407135
CCN	Koontz, A., Uin, J., Andrews, E., Enekwizu, O., Hayes, C., & Salwen, C. Cloud Condensation Nuclei Particle Counter (AOSCCN2COLA), 2023-02-15 to 2024-02-14, ARM Mobile Facility (EPC), La Jolla, CA; AMF1 (main site for EPCAPE on Scripps Pier) (M1). Atmospheric Radiation Measurement (ARM) User Facility. https://doi.org/10.5439/1323892
CO-ANALYZER	Sedlacek, A., Hayes, C., Koontz, A., Springston, S., & Trojanowski, R. Carbon Monoxide Analyzer (AOSCO), 2023-02-15 to 2024-02-14, ARM Mobile Facility (EPC), La Jolla, CA; AMF1 (main site for EPCAPE on Scripps Pier) (M1). Atmospheric Radiation Measurement (ARM) User Facility. https://doi.org/10.5439/1250819
CPC	Singh, A., Flynn, C., Ermold, B., Andrews, E., Kuang, C., Hayes, C., & Salwen, C. Condensation Particle Counter (AOSPCF1M), 2023-02-15 to 2024-02-14, ARM Mobile Facility (EPC), La Jolla, CA; AMF1 (main site for EPCAPE on Scripps Pier) (M1). Atmospheric Radiation Measurement (ARM) User Facility. https://doi.org/10.5439/1398277
HTDMA	Uin, J., Cromwell, E., Hayes, C., & Salwen, C. Humidified Tandem Differential Mobility Analyzer (AOSHTDMA), 2023-02-15 to 2024-02-14, ARM Mobile Facility (EPC), La Jolla, CA; AMF1 (main site for EPCAPE on Scripps Pier) (M1). Atmospheric Radiation Measurement (ARM) User Facility. https://doi.org/10.5439/1095581
NEPHELOMETER	Uin, J., Salwen, C., Senum, G., Hayes, C., & Mayol-Bracero, O. L. Nephelometer (AOSNEPHDRY1M), 2023-02-15 to 2024-02-14, ARM Mobile Facility (EPC), La Jolla, CA; AMF1 (main site for EPCAPE on Scripps Pier) (M1). Atmospheric Radiation Measurement (ARM) User Facility. https://doi.org/10.5439/1984586
OZONE	Sedlacek, A., Hayes, C., Springston, S., Koontz, A., & Trojanowski, R. Ozone Monitor (AOSO3), 2023-02-15 to 2024-02-14, ARM Mobile Facility (EPC), La Jolla, CA; AMF1 (main site for EPCAPE on Scripps Pier) (M1). Atmospheric Radiation Measurement (ARM) User Facility. https://doi.org/10.5439/1346692
PSAP	Sedlacek, A., Smith, S., Ermold, B., Flynn, C., Trojanowski, R., Andrews, E., Hayes, C., Salwen, C., & Flynn, C. Particle Soot Absorption Photometer (AOSPSAP3W1M), 2023-02-15 to 2024-02-14, ARM Mobile Facility (EPC), La Jolla, CA; AMF1 (main site for EPCAPE on Scripps Pier) (M1). Atmospheric Radiation Measurement (ARM) User Facility. https://doi.org/10.5439/1225037
SMPS	Singh, A., Kuang, C., Howie, J., Salwen, C., & Hayes, C. Scanning mobility particle sizer (AOSMPS), 2023-02-15 to 2024-02-14, ARM Mobile Facility (EPC), La Jolla, CA; AMF1 (main site for EPCAPE on Scripps Pier) (M1). Atmospheric Radiation Measurement (ARM) User Facility. https://doi.org/10.5439/1476898
SO2	Sedlacek, A., Hayes, C., & Trojanowski, R. Sulfur Dioxide Monitor (AOSSO2), 2023-02-15 to 2024-02-14, ARM Mobile Facility (EPC), La Jolla, CA; AMF1 (main site for EPCAPE on Scripps Pier) (M1). Atmospheric Radiation Measurement (ARM) User Facility. https://doi.org/10.5439/1250820
SP2	Jackson, R., Sedlacek, A., Salwen, C., & Hayes, C. Single Particle Soot Photometer (AOSSP2BC60S), 2023-02-15 to 2024-02-14, ARM Mobile Facility (EPC), La Jolla, CA; AMF1 (main site for EPCAPE on Scripps Pier) (M1). Atmospheric Radiation Measurement (ARM) User Facility. https://doi.org/10.5439/1807910

ARM Data Object Identifier (DOI) Table

Instrument	DOI
UHSAS	Uin, J., Senum, G., Koontz, A., Flynn, C., Salwen, C., & Hayes, C. Ultra-High Sensitivity Aerosol Spectrometer (AOSUHSAS), 2023-02-15 to 2024-02-14, ARM Mobile Facility (EPC), La Jolla, CA; AMF1 (main site for EPCAPE on Scripps Pier) (M1). Atmospheric Radiation Measurement (ARM) User Facility. https://doi.org/10.5439/1409033