

# Absolute calibration of gravitational wave detector using gravity field and photon pressure

Yuki Inoue[1], Sadakazu Haino, Nobuyuki Kanda, Yujiro Ogawa, Toshikazu Suzuki, Takayuki Tomaru, Takahiro Yamamoto, Takaaki Yokozawa

Institute of Physics, Academia Sinica, Nankang, Taipei 11529, Taiwan

Institute for Cosmic Ray Research, University of Tokyo, Kashiwa, Chiba, 277-8582, Japan

Department of Physics, Osaka City University, Sumiyoshi, Osaka 558-8585, Japan

KAGRA Observatory, Institute for Cosmic Ray Research, University of Tokyo, Hida, Gifu 506-1205, Japan

High Energy Accelerator Research Organization, Tsukuba, Ibaraki, 305-0801, Japan

E-mail: iyuki@post.kek.jp

June 2018

**Abstract.** Absolute calibration of the gravitational wave detectors are essential for understanding of source parameters of the gravitational wave source. Current limit of the uncertainty of the absolute laser calibration is XXXX % corresponding to the absolute laser power of the photon calibrator. The photon calibrator is primary calibrator and calibrate the absolute displacement of the mirror by using the photon pressure. In order to reduce the uncertainty of photon calibrator, we propose the use of gravity field calibrator. We can modulate the mirror with a precision less than XXXX % when we assume the basic parameter of KAGRA.

## 1. Introduction

The discovery of the gravitational wave give us the new probe for observing our universe. Instruments for measuring the gravitational wave must be designed for accuracy as well as precession. The typical strain sensitivity of 2nd generation interferometric detectors(IFO), such as Advanced LIGO, Advanced Virgo, and KAGRA, are around  $10^{-23}/\sqrt{Hz}$ . In order to estimate the parameters of gravitational waves, such as masses, spins, redshift and distance, understanding of the systematic and statistical noise sources are critical.

To reduce the systematic errors of the gravitational wave calibration and reconstruction, we need to calibrate the response of IFO. The first generation of the photon calibrator is developed by the Gragrow. LIGO and Virgo employ the second generation photon calibrator for the calibration of time-dependent optical gain of IFO. The frequency range of photon calibrator is between 10 and 3000Hz. However, it has

a challenging issue of the absolute calibration due to the accuracy of the absolute laser power of laser standard between each country. Therefore, current uncertainty of absolute power is a few percent.

The dynamic gravity field generator is one of the candidates to be able to solve the uncertainty problem of absolute calibration. Related techniques using photon calibrator and gravity field calibrator are described as Karuki. et. al and Matone et al.

## 2. Photon calibrator


Photon calibrator (Pcal) relies on the photon radiation pressure from the power modulated laser beams reflecting on the test mass to apply periodic force via the recoil of photons. LIGO, Virgo and KAGRA employ the photon calibrator for the calibration of the interferometer response. The displacement of the test mass can be described as

$$x = \frac{P(\omega) \cos \theta}{2c} s(\omega) \left( 1 + \frac{M}{I} \vec{a} \cdot \vec{b} \right), \quad (1)$$

where  $P$  is absolute laser power,  $\theta$  is incident angle of the Pcal laser,  $M$  is mass of test mass,  $\omega$  is angular frequency,  $\vec{a}$  and  $\vec{b}$  are position vector of Pcal laser beams.  $I = Mh^2/12 + Mr^2/4$  is inertia, where  $h$  and  $r$  are thickness and radius of test mass.  $s(\omega)$  is transfer function between force and displacements. We can regard the  $s(\omega)$  as  $1/(M\omega^2)$  above 20 Hz. The laser power is stabilized less than design sensitivity. The stabilized laser is mounted on the transmitter module. The power is monitored by the average response of the photo detectors at the transmitter module and receiver module. The uncertainty of displacement corresponds to uncertainty of laser power. The largest uncertainty of photon calibrator is uncertainty of laser power. LIGO and KAGRA use the working standard to cross-calibrate the relative response of each interferometer. The uncertainty of each calibrator is 0.51 %. For the absolute calibration, it is compared with the NIST laser power standard. We send Gold standard to NIST and calibrate the response of detector. After that, we bring the gold standard to LHO and calibrate the working standard Hanford, Livingston and KAGRA. The uncertainty of laser standard of each institute of standard is a few percent. The second largest uncertainty of photon calibrator is optical efficiency of optical path. the estimated uncertainty of optical efficiency in LIGO is 0.37 %. We need to estimate the laser power response through the response of photo detector at the outside of vacuum chamber. The reduction of these uncertainties are essential for the gravitational wave observation.

## 3. Gravity field calibrator

To solve the calibration uncertainty, we propose the gravity field calibrator. The original idea is described in Matone. et.al. The gravity field calibrator generate the gravity field at the end of test mass. By rotating the multipole masses, the rotor placed in the vacuum chamber for isolating the acoustic noise. To monitor the frequency, we mount the 10-bit encoder. We monitor the frequency using 16 bit ADC system. We calculated



**Figure 1.** Schematic view of photon calibrator. We place the stabilized laser on the transmitter module. The injected signal at the test masses is monitored by using the average of photo detector power between the transmitter module,  $V_{TxPD}$  and receiver module,  $V_{RxPD}$ . The geometrical factor is characterized by the position vector of the main beam,  $\vec{a} = \vec{a}_1 + \vec{a}_2$ , and photon calibrator beams,  $\vec{b}$ .

**Table 1.** Specification summary of LIGO, Virgo and KAGRA photon calibrator


	KAGRA	advanced LIGO	advanced Virgo
Mirror material	Sapphire	Silica	Silica
Mirror mass	22.8 kg	40 kg	40 kg
Mirror diameter	220 mm	340 mm	350 mm
Mirror thickness	150 mm	200 mm	200 mm
Distance of Pcal from ETM	36 m	8 m	1.5 m
Pcal laser power	20 W	2W	3 W
Laser frequency	1047 nm	1047 nm	1047 nm
Incident angle	0.72 deg	8.75 deg	30 deg

the displacement by changing dynamic gravity field of multipole moment with with  $N$  masses. The assumed model is the suspended test masses for the interferometer and disc with multipole mass as shown in Fig ???. We put the masses  $m$  at the positions of the radius  $r$ . The distance between the center of mass of mirror and disc is assumed  $d$ . We rotate the disc at the angular frequency of  $\omega_{rot} = 2\pi f_{rot}$ .

Distance between  $i$ -th mass and center of test mass is

$$L_i = d \sqrt{1 + \left(\frac{r}{d}\right)^2 - 2 \left(\frac{r}{d}\right) \cos \phi_i}, \quad (2)$$

where the angle of  $i$ -th mass is assumed as  $\phi_i = \omega_{rot}t + 2\pi i/N$ . The gravity potential



**Figure 2.** Schematic view of gravity field calibrator. We placed the rotor at the displacement of  $d$  away from test masses. Multipole mass generate the gravitational potential at the test mass position.

at the center of test mass can be described as

$$V = \sum_{i=0}^N V_i \quad (3)$$

$$= -\frac{GMm}{d} \sum_{i=0}^N \sum_{n=0}^{\infty} \left(\frac{r}{d}\right)^n P_n \left(\cos \left(\omega_{rot}t + \frac{2\pi i}{N}\right)\right), \quad (4)$$

where  $P_n$  is Legendre polynomial. The equation of motion of test mass is

$$Ma = \left| \frac{\partial V}{\partial d} \right| = \frac{GMm}{d^2} \sum_{i=0}^N \sum_{n=0}^{\infty} (n+1) \left(\frac{r}{d}\right)^n P_n \left(\cos \left(\omega_{rot}t + \frac{2\pi i}{N}\right)\right), \quad (5)$$

where  $a$  is acceleration of test mass. We place the quadrupole and hexapole masses in same rotor as shown in Fig. ???. We put the hole between each mass. The hole can increase the gravity gradient twice at the test massis. Therefore, we can describe the equation of motion as


$$Ma = \left| \frac{\partial V}{\partial d} \right| = \frac{2GMm}{d^2} \sum_{i=0}^N \sum_{n=0}^{\infty} (n+1) \left(\frac{r}{d}\right)^n P_n \left(\cos \left(\omega_{rot}t + \frac{2\pi i}{N}\right)\right), \quad (6)$$

We will calculate the displacement of quadrupole and hexapole in the section ??

### 3.1. Displacement of test mass(Quadrupole)

We calculate the displacement of the quadrupole mass distribution, which corresponds to  $N = 2$ . The mass and radius of quadrupole are assumed as  $m_q$ ,  $r_q$ . The equation of motion of test mass is described as

$$Ma = \frac{2GMm_q}{d^2} \sum_{n=0}^{\infty} (n+1) \left(\frac{r_q}{d}\right)^n (P_n(\cos(\omega_{rot}t)) + P_n(\cos(\omega_{rot}t + \pi))). \quad (7)$$



**Figure 3.** Configuration of the rotor with quadrupole and hexapole mass distribution.  $m_q$  and  $m_h$  are mass of quadrupole and hexapole.  $r_q$  and  $r_h$  are radius of quadrupole and hexapole.

If we assume  $r \ll d$ , the displacement of the time-dependent lower harmonics can be written by

$$x = \sum_{i=1}^{\infty} x_{ni} \cos(n\omega_{rot}t) \sim x_{2f} \cos(2\omega_{rot}t) = x_{2f} \cos \omega t, \quad (8)$$

where amplitude of 2-f rotation is

$$x_{2f} = 9 \frac{GMm_q r_q^2}{d^4} s(\omega). \quad (9)$$

### 3.2. Displacement of test mass(Haxapole)

We calculate the displacement of the hexapole mass distribution, which corresponds to  $N = 3$ . The mass and radius of hexapole are assumed as  $m_h$ ,  $r_h$ . The equation of motion of test mass is described as

$$Ma = \frac{2GMm_h}{d^2} \sum_{n=0}^{\infty} (n+1) \left( \frac{r_h}{d} \right)^n \quad (10)$$

$$\times \left( P_n(\cos(\omega_{rot}t)) + P_n\left(\cos\left(\omega_{rot}t + \frac{2\pi}{3}\right)\right) + P_n\left(\cos\left(\omega_{rot}t + \frac{4\pi}{3}\right)\right) \right). \quad (11)$$

If we assume  $r \ll d$ , the displacement of the time-dependent lower harmonics can be written by

$$x = \sum_{i=1}^{\infty} x_{ni} \cos(n\omega_{rot}t) \sim x_{3f} \cos(3\omega_{rot}t) = x_{3f} \cos \omega t, \quad (12)$$

where amplitude of 3-f is

$$x_{3f} = 15 \frac{GMm_h r_h^3}{d^5} s(\omega). \quad (13)$$

#### 4. Absolute power calibration by using Gcal and Pcal

In this section, we will discuss about absolute laser power calibration using interferometer. Figure ?? shows the configuration of absolute laser power calibration by using the combination of photon calibrator and gravity field calibrator. First, we modulate the mirror position using gravity field calibrator. We can measure the signal of  $x_{2f}$  and  $x_{3f}$  in the response of interferometer. Second, we send the interferometer signal to the excitation port of photon calibrator as a reference of feedback control. The photon calibrator cancel the displacement by gravity field calibrator. Third, we measure the absolute laser power at the detector of transmitter module and receiver module. The output signal of transmitter module,  $V_{TxPD}$  and receiver module,  $V_{RxPD}$  corresponding to displacement from gravity field should be modulated. By using Eq (??), (??), and (??), the modulated powers are

$$P_{2f} = 18 \frac{Gm_q Mr_q^2}{d^4 \cos \theta} \frac{1}{1 + \frac{M}{I} \vec{a} \cdot \vec{b}} \quad (14)$$

$$P_{3f} = 30 \frac{Gm_h Mr_h^3}{d^5 \cos \theta} \frac{1}{1 + \frac{M}{I} \vec{a} \cdot \vec{b}} \quad (15)$$

Fourth, we demodulate the signal of transmitter module and receiver module using the encoder signal of gravity field calibrator. The demodulated signals are

$$V_{2f}^T = \rho_T P_{2f} \quad (16)$$

$$V_{2f}^R = \rho_R P_{2f} \quad (17)$$

$$V_{3f}^T = \rho_T P_{3f} \quad (18)$$

$$V_{3f}^R = \rho_R P_{3f} \quad (19)$$

Therefore, we can estimate the distance using measured data.

$$d = \frac{5 V_{2f}^T m_h r_h^3}{3 V_{3f}^T m_q r_q^2} = \frac{5 V_{2f}^R m_h r_h^3}{3 V_{3f}^R m_q r_q^2} \quad (20)$$

Using Eq (??), (??) and (??), we can calibrate the absolute laser power. Finally, we insert the equation (??) to the equation (??):


$$x = \frac{P(\omega) \cos \theta}{2c} s(\omega) \left( 1 + \frac{M}{I} \vec{a} \cdot \vec{b} \right) \quad (21)$$

$$= 9 \frac{P(\omega)}{P_{2f}} \frac{Gm_q Mr_q^2}{d^4} s(\omega), \quad (22)$$

$$= \frac{729}{625} \frac{Gm_q^5 r_q^{14}}{m_h^4 r_h^{12} \omega^2} \frac{V_{3f}^{R4}}{V_{2f}^{R5}} V^R(\omega). \quad (23)$$

$$= \frac{729}{625} \frac{Gm_q^5 r_q^{14}}{m_h^4 r_h^{12} \omega^2} \frac{V_{3f}^{R4}}{V_{2f}^{R5}} V_0^R \cos \omega t. \quad (24)$$

$$(25)$$



**Figure 4.** Test setup of the calibration of the laser power by rotating gravity field calibrator.

We estimate the relative uncertainty of displacement

$$\left(\frac{\delta x}{x}\right)^2 \sim \left(\frac{\delta G}{G}\right)^2 + \left(\frac{\delta V_0^R}{V_0^R}\right)^2 + 4\left(\frac{\delta \omega}{\omega}\right)^2 + 25\left(\frac{\delta m_q}{m_q}\right)^2 + 16\left(\frac{\delta m_h}{m_h}\right)^2 \quad (26)$$

$$+ 25\left(\frac{\delta V_{f2}^R}{V_{f2}^R}\right)^2 + 16\left(\frac{\delta V_{f3}^R}{V_{f3}^R}\right)^2 + 196\left(\frac{\delta r_q}{r_q}\right)^2 + 144\left(\frac{\delta r_h}{r_h}\right)^2 \quad (27)$$

## 5. Estimation of uncertainty

We estimate the uncertainty of laser power by assuming KAGRA. The assumed parameters are listed in Table XXX. The estimated RMS of displacement are

$$x_{2f}^{rms} = \quad (28)$$


$$x_{3f}^{rms} = \quad (29)$$

The expected powers are

$$P_{2f} = \quad (30)$$

$$P_{3f} = \quad (31)$$

The calculated  $V_{2f}^T/V_{3f}^T$  by changing distance are shown in Fig.XXX. We assumed the gaussian noise for each the parameters. We estimate the uncertainty with Eq XXXX. Table XXXX shows the uncertainty of photon calibrator system. The uncertainty of the quadrupole and hexapole mass is limited by the accuracy of electronic balance. In this case, we assumed that we use CG-6000. The accuracy of CG-6000 is 0.2 g. When we make the disk, we use the NC milling. Its accuracy is 0.02 mm. We can measure the dimension with three-dimension coordinate measuring machine (CMM). The precision of



**Figure 5.**

**Table 2.** The assumed parameters.

		Value	Relative uncertainty
$G$	Gravity constant	6.6742	0.015 %
$c$	speed of light[m/sec]	2.99792458e8	
$I$	Inertia	0.1127[kgm <sup>2</sup> ]	0.6 %
$\theta$	Incident andgle	0.7[deg]	0.07%
$M$	Mass of test mass[kg]	23	0.02 %
$m_q$	Mass of quadrupole[kg]	4.485	0.004 %
$m_h$	Mass of hexapole[kg]	4.485	0.004%
$r_q$	Radius of quadrupole[m]	0.200	0.025 %
$r_h$	Radius of hexapole[m]	0.125	0.04%
$\frac{I}{M} \vec{a} \cdot \vec{b}$	Geometrical factor		0.3 %

CMM is 2  $\mu\text{m}$ . We can measure the position of mass accurately. By using Gcal system, we can reduce the uncertainty less than XXXX.

## 6. Discussion

This method has an advantage of a direct comparison of the amplitude of injected power and gravity field. In previous study, we need to consider the uncertainty of the optical efficiency through the window and mirrors. This is because we put the working standard at the outside of the chamber. However, the method of gravity field can compare the displacement directly. By using this method, we can calibrate the uncertainty of optical efficiency and absolute power of the laser. When we compare the laser power between each institute, we need to bring the working standard. However, we



**Table 3.**  $N=2, \omega = n\omega_{rot}$ 

modulation	n=1	n=2	n=3	n=4	n=5	n=6
1f	0	0	0	0	0	0
2f	0	$9 \frac{Gmr^2}{d^4\omega^2}$	0	$\frac{25}{4} \frac{Gmr^4}{d^6\omega^2}$	0	$\frac{33075}{128} \frac{Gmr^6}{d^8\omega^2}$
3f	0	0	0	0	0	0
4f	0	0	0	$\frac{175}{16} \frac{Gmr^4}{d^6\omega^2}$	0	$\frac{3675}{64} \frac{Gmr^6}{d^8\omega^2}$
5f	0	0	0	0	0	0
6f	0	0	0	0	0	$\frac{1617}{128} \frac{Gmr^6}{d^8\omega^2}$

**Table 4.**  $N=3, \omega = n\omega_{rot}$ 

modulation	n=1	n=2	n=3	n=4	n=5	n=6
1f	0	0	0	0	0	0
2f	0	0	0	0	0	0
3f	0	0	$15 \frac{Gmr^3}{d^5\omega^2}$	0	0	$\frac{1449}{32} \frac{Gmr^5}{d^7\omega^2}$
4f	0	0	0	0	0	0
5f	0	0	0	0	0	0
6f	0	0	0	0	0	$\frac{4851}{256} \frac{Gmr^6}{d^8\omega^2}$

can try the absolute calibration using this method. As we mention about a few percent of the absolute uncertainty of laser in each country. The estimated uncertainty of the power of this method is less than 1 %. It imply that we can make a new power standard using interferometer.

The optimization of the parameter of the rotor is important. To reduce the uncertainty of displacement, we put the quadrupole at the outer side and hexapole at the inner side. The estimated the distance uncertainty is XXXXX. If we replace the position between quadrupole and hexapole, it becomes XXX. The uncertainty of the radius propagate to the uncertainty of distance and resolution of the distance also increse.

We need to pay attention to the modulation effect of the intermediate mass because the gravity field act all the mass. We calculate the responses of the each masses using transfer function of KAGRA as shown in Fig XXX. We can neglect the intermediate mass effect and regard as free mass motion larger than XXXXX Hz.

If we consider the precision less than 1 %, we need to consider the higher order of Legendre polynomial. This is because that higher order also has 2-f and 3-f components. These factor is mitigated by the factor of  $(r/g)^n$ , where  $n$  is order of Legendre polynomial. Table XXX shows the higher order term. To compare the higher order effect by changing distance, we simulate the  $x_{2f}$  and  $x_{3f}$  using the FET method.

## 7. Conclusion

Photon calibrator is one of the powerful calibrators in LIGO, Virgo and KAGRA. It can calibrate the response of IFO and its uncertainty is essential for estimation of gravitational wave source. In particular, the distance of the source is strongly depend on the absolute laser power of the photon calibrator. In previous study, the Gold standard, which response is calibrated by the laser power standard of NIST, is used for the absolute laser power calibration of the photon calibrator. However, current limit of the absolute laser power between each country is about XXX %. It is directly propagate to the uncertainty of absolute displacement of gravitational wave detector.

To solve the problem, we proposed the combination method of photon calibrator and gravity field calibrator. Gravity field calibrator can modulate the mirror using gravity gradient. By canceling the displacement of the test mass using the photon calibrator, we can calibrate the absolute laser power and displacement of the photon calibrator with accuracy of 0.XXX %. Improvement of the uncertainty of absolute calibration is essential for parameter estimation.