
A BV simulator vs Laviron (1979)
BV-SWV E_{sw} = 25 mV (operating)
BV-SWV E_{sw} = 1 mV (small-E_{sw} limit)
Laviron (1979) analytic
FWHM = 60.6 mV
ΔI (normalised)
Potential (mV)
B Laviron peak: FWHM = 3.53 $RT/(n_eF)$
n_e = 1, FWHM = 90.7 mV
n_e = 2, FWHM = 45.3 mV
n_e = 4, FWHM = 22.7 mV
|Ψ| (normalised)
Overpotential η (mV)
C Frequency dependence (DD2017 parameters)
f = 200 Hz, κ = 12.5, FWHM = 56 mV
f = 500 Hz, κ = 5.0, FWHM = 60 mV
f = 1000 Hz, κ = 2.5, FWHM = 62 mV
ΔI (normalised)
Potential (mV)
D FWHM vs κ
Laviron limit 45.3 mV
$κ_{tip}$ = 6.4, FWHM = 59 mV
FWHM (mV)
$κ = k^0/(2f)$