
1. Scan before
SWV scan of the fresh electrode: peak height $\Delta I_{MB}(0)$
2. Incubate
wound fluid sits on the electrode for time T; MMP-9 cleaves probes
3. Scan after
same electrode, same settings: peak height $\Delta I_{MB}(T)$
4. Read out
$F_k = 1 - \frac{\Delta I_{MB}(T)}{\Delta I_{MB}(0)}$
then $F_k \rightarrow$ [MMP-9]