DENTAL LABORATORY SCANNER

LS2

A practical start to digital scanning.



Everyday model digitisation

11 s

Full-arch scan

2 × 2.4 MP

Camera configuration

6 µm* | 9 kg

Stated accuracy / Weight

01

11-second full-arch scan

A listed full-arch scan time of 11 seconds, with 27 seconds for a full-arch impression.

02

Dual-camera capture

Two 2.4 MP cameras combine with blue-light scanning and true-colour texture capture.

03

Open digital output

Export STL, PLY or OBJ files to continue working in compatible CAD software.

Technical specifications and computer requirements on the reverse. * See LS2 accuracy note in the specifications.

TECHNICAL SPECIFICATIONS

LS2 at a glance.

Camera configuration	2 × 2.4 MP
Accuracy (ISO 12836)	6 µm*
Full-arch scan	11 s
Full-arch impression scan	27 s
Scan area	60 × 84 × 106 mm
Scanner dimensions	393 × 271 × 274 mm
Scanner weight	9 kg
Package dimensions	40 × 40 × 50 cm
Output formats	STL / PLY / OBJ
Light source	Blue light
Texture scanning	True colour
Power supply	DC 24 V
Interface	USB 3.0
Articulator scanning	Dynamic / static; articulator transfer

COMPUTER CONFIGURATION

Component	Minimum	Recommended
Processor	Intel Core i7-10700 or above	Intel Core i7-13700 or above
Memory	16 GB or above	32 GB or above
Graphics	NVIDIA GeForce GTX 1660 6 GB or above	NVIDIA GeForce RTX 2060 SUPER 8 GB or above
Operating system	Windows 10 (64-bit)	Windows 11 (64-bit)
Connection	USB 3.0	USB 3.0

Source: supplied Spectra Scanner catalogue, pp. 2, 4-5. Supplier-stated figures; scan times may vary. Dimension order follows the source. Confirm current OS support before installation. * LS2: the source prints the inequality after the 6 µm value. The numeric value is retained; the direction of the limit requires supplier confirmation.