



Chemiluminescence Immunology Analyzer Installation Report

RCLIA-24001-All products(V1.0)

Customer Name		Installation Order		
Analyzer Model	CL-900i	Analyzer SN		
Engineer Name		Installation Date		
Basic Acceptance				
No.	Chemistry	Content prompt	Conclusion	Remarks
1	Acceptance of Appearance, Quantity, Etc	1) Internal and external packaging (no damage, bumps, soaking, damp, deformation, etc.) 2) Appearance of equipment and accessories (without damage, rust, bumps, etc.) 3) The model, specification, configuration and quantity of the main unit and accessories comply with the packing list. 4) The accompanying documents (operation instructions, maintenance card, product certificate, etc.) are complete.		If the appearance is damaged, please indicate the amount and degree of damage. If the number of accessories is abnormal, please indicate the name and quantity of the materials unqualified:
2	Working Environment	1) Ambient temperature (15~30°C): _____; 2) Humidity(35~85%): _____; 3) Voltage(V): _____; 4) Ground voltage(<5V): _____; 5) Altitude(m): _____;		
3	System Status Check	The middle-layer unit displays OK, without exception in red All subunits of the slave are OK without exception in red 1) The temperature option is not marked in red. 2) Power options are not marked in red 3) No exception is marked in red for fluidic options.		If there are unqualified items, please check the hardware connection and tubing connection of the instrument and make some adjustments
4	Motion Positioning Performance Test	Sample Probe unit Reagent probe unit Gripper Unit Dispersion Unit Reaction Carousel/Incubation Unit Reagent carousel unit Waste Aspirating Unit Cuvette Loader Unit Sample Delivery Unit		If there is any unqualified item, please use the alignment tool to calibrate it and keep the parameter records. (Unqualified item - original parameter - Adjusted parameter) _____:_____:_____; _____:_____:_____; _____:_____:____;
5	SPT	DCF Stability Test (R.E.D< 1.5%) : _____; Dark Count (Maximum< 1.5%) : _____; Substrate background test (SD< 100) : _____; Sample Probe System Repeatability (CV<1.5%)_____; Reagent Probe System Repeatability (CV<1.5%)_____;		If there are unqualified items, please check whether the substrate and its tubes are contaminated, and whether the sampling verticality and horizontal position are appropriate.
6	Bar Code Performance Test	Reagent Bar Code Scanning Stability Test Sample Bar Code Stability Test		If no passed, please check if the barcode meets the requirements and if the scanning position is appropriate.
Overall conclusion of installation acceptance				
Exception Description and Suggestions				