

System Log Report: Warning and Errors

From: 12/1/2020

To: 1/4/2021

<u>Log Date</u>	<u>Message</u>	<u>Code</u>	<u>Log Type</u>	<u>Error Level</u>	<u>User ID</u>
12/6/2020 3:10:09AM	The Luminex Detector Device is not connected or turned off.	100CB00E	System	W	None
12/6/2020 3:10:09AM	The XY Device is not connected or turned off.	100CB00F	System	W	None
12/13/2020 3:10:17AM	The Luminex Detector Device is not connected or turned off.	100CB00E	System	W	None
12/13/2020 3:10:17AM	The XY Device is not connected or turned off.	100CB00F	System	W	None
12/16/2020 4:51:04PM	The Luminex Detector Device is not connected or turned off.	100CB00E	System	W	admin
12/16/2020 4:51:04PM	The XY Device is not connected or turned off.	100CB00F	System	W	admin
12/16/2020 4:51:04PM	The Luminex Detector Device is not connected or turned off.	100CB00E	System	W	admin
12/16/2020 4:51:04PM	The XY Device is not connected or turned off.	100CB00F	System	W	admin
12/16/2020 4:52:30PM	The Luminex Detector Device is not connected or turned off.	100CB00E	System	W	admin
12/16/2020 4:52:51PM	The Luminex Detector Device is not connected or turned off.	100CB00E	System	W	None
12/16/2020 4:52:51PM	The XY Device is not connected or turned off.	100CB00F	System	W	None
12/16/2020 4:52:51PM	The Luminex Detector Device is not connected or turned off.	100CB00E	System	W	None
12/16/2020 4:52:51PM	The XY Device is not connected or turned off.	100CB00F	System	W	None
12/30/2020 4:38:13PM	The Luminex Detector Device is not connected or turned off.	100CB00E	System	W	None
12/30/2020 4:38:13PM	The XY Device is not connected or turned off.	100CB00F	System	W	None
12/30/2020 4:38:22PM	The Luminex Detector Device is not connected or turned off.	100CB00E	System	W	None
12/30/2020 5:09:32PM	The Luminex Detector Device is not connected or turned off.	100CB00E	System	W	None
12/30/2020 5:09:32PM	The XY Device is not connected or turned off.	100CB00F	System	W	None
12/30/2020 5:09:32PM	The Luminex Detector Device is not connected or turned off.	100CB00E	System	W	None

System Log Report: Warning and Errors

From: 12/1/2020

To: 1/4/2021

12/30/2020 5:09:32PM	The XY Device is not connected or turned off.	100CB00F	System	W	None
12/30/2020 5:10:49PM	The Luminex Detector Device is not connected or turned off.	100CB00E	System	W	
12/30/2020 5:11:15PM	The Luminex Detector Device is not connected or turned off.	100CB00E	System	W	None
12/30/2020 5:11:15PM	The XY Device is not connected or turned off.	100CB00F	System	W	None
12/30/2020 5:11:15PM	The Luminex Detector Device is not connected or turned off.	100CB00E	System	W	None
12/30/2020 5:11:15PM	The XY Device is not connected or turned off.	100CB00F	System	W	None
12/31/2020 2:07:31PM	The XY Device is not connected or turned off.	100CB00F	System	W	admin
12/31/2020 2:07:51PM	The XY Device is not connected or turned off.	100CB00F	System	W	admin
12/31/2020 2:54:23PM	Detector Device has gone into an error state. The error was: The device has timed out during an operation and needs to be reset. (0xCFF01037)	CFF01037	Maintenance	E	admin
12/31/2020 3:07:31PM	The Luminex Detector Device is not connected or turned off.	100CB00E	System	W	admin
12/31/2020 3:07:31PM	The XY Device is not connected or turned off.	100CB00F	System	W	admin
12/31/2020 3:07:31PM	The Luminex Detector Device is not connected or turned off.	100CB00E	System	W	admin
12/31/2020 3:07:31PM	The XY Device is not connected or turned off.	100CB00F	System	W	admin
12/31/2020 3:24:15PM	The Luminex Detector Device is not connected or turned off.	100CB00E	System	W	None
12/31/2020 3:24:15PM	The XY Device is not connected or turned off.	100CB00F	System	W	None
12/31/2020 3:25:21PM	The Luminex Detector Device is not connected or turned off.	100CB00E	System	W	admin
12/31/2020 3:26:35PM	Warmup command completed with status FAILED. The command was cancelled with an abort. (0xCFF01093)	CFF01093	Maintenance	E	admin
12/31/2020 3:26:35PM	Warmup_None command completed with status FAILED. The command was cancelled with an abort. (0xCFF01093)	CFF01093	Maintenance	E	admin
12/31/2020 3:26:36PM	Routine 'Daily Instrument Startup AMP (Luminex)' aborted.	800C9638	Maintenance	E	admin

System Log Report: Warning and Errors

From: 12/1/2020 To: 1/4/2021

12/31/2020 3:31:26PM	The Luminex Detector Device is not connected or turned off.	100CB00E	System	W	admin
12/31/2020 3:31:26PM	The XY Device is not connected or turned off.	100CB00F	System	W	admin
12/31/2020 3:31:45PM	The Luminex Detector Device is not connected or turned off.	100CB00E	System	W	admin
12/31/2020 3:31:45PM	The XY Device is not connected or turned off.	100CB00F	System	W	admin
12/31/2020 3:37:07PM	Move To (0,B7) command completed with status FAILED. The command could not be sent to the device. (0xCFF0104E)	800C9580	Maintenance	E	admin
12/31/2020 3:37:07PM	Wash_1,B7 command completed with status FAILED. The command could not be sent to the device. (0xCFF0104E)	CFF0104E	Maintenance	E	admin
12/31/2020 3:38:11PM	The Luminex Detector Device is not connected or turned off.	100CB00E	System	W	admin
12/31/2020 3:38:11PM	The XY Device is not connected or turned off.	100CB00F	System	W	admin
12/31/2020 3:41:43PM	The Luminex Detector Device is not connected or turned off.	100CB00E	System	W	admin
12/31/2020 3:41:43PM	The XY Device is not connected or turned off.	100CB00F	System	W	admin
12/31/2020 3:49:17PM	Warmup command completed with status SUCCEEDED. The command was cancelled. (0x4FF01052)	4FF01052	System	W	admin
12/31/2020 3:49:17PM	User cancelled warmup.	100CB007	Maintenance	W	admin
12/31/2020 3:49:56PM	Detector Device has gone into an error state. The error was: The device has timed out during an operation and needs to be reset. (0xCFF01037)	CFF01037	Maintenance	E	admin
12/31/2020 4:09:35PM	The Luminex Detector Device is not connected or turned off.	100CB00E	System	W	admin
12/31/2020 4:09:35PM	The XY Device is not connected or turned off.	100CB00F	System	W	admin
12/31/2020 4:09:35PM	The Luminex Detector Device is not connected or turned off.	100CB00E	System	W	admin
12/31/2020 4:09:35PM	The XY Device is not connected or turned off.	100CB00F	System	W	admin
1/3/2021 9:11:53AM	Detector Device has gone into an error state. The error was: The device has timed out during an operation and needs to be reset. (0xCFF01037)	CFF01037	Maintenance	E	admin

System Log Report: Warning and Errors

From: 12/1/2020 To: 1/4/2021

1/3/2021	9:11:53AM	Calibration Classification (CAL1) command completed with status FAILED. The device has timed out during an operation and needs to be reset. (0xCFF01037)	800C9580	Maintenance	E	admin
1/3/2021	9:11:53AM	Plate command completed with status FAILED. The device has timed out during an operation and needs to be reset. (0xCFF01037)	CFF01037	Maintenance	E	admin
1/3/2021	9:11:54AM	Maintenance controller error on End Well	800C9580	Maintenance	E	admin
1/3/2021	9:11:54AM	The added or subtracted value results in an un-representable DateTime.Parameter name: value	800C9580	Maintenance	E	admin
1/3/2021	9:11:54AM	Maintenance controller timeout error on End Plate	800C9613	Maintenance	E	admin
1/3/2021	9:11:55AM	The XY Device is not connected or turned off.	100CB00F	System	W	admin
1/3/2021	9:12:03AM	The XY Device is not connected or turned off.	100CB00F	System	W	admin
1/3/2021	11:38:08AM	The Luminex Detector Device is not connected or turned off.	100CB00E	System	W	admin
1/3/2021	11:38:08AM	The XY Device is not connected or turned off.	100CB00F	System	W	admin
1/3/2021	11:38:08AM	The Luminex Detector Device is not connected or turned off.	100CB00E	System	W	admin
1/3/2021	11:38:08AM	The XY Device is not connected or turned off.	100CB00F	System	W	admin