

OPERATION QUALIFICATION (OQ)

Operation Qualification (OQ) testing is performed to verify that the instrument or instrument modules operate as intended.

Instrument:	DG SPIN (DG-320)	Form version:	1.0
Serial number:		Code:	
Installation date:			

Site:	Operator Name:		
Department:			
Tel:	Fax:	E-Mail:	

1 Connect:

- ☐ DG SPIN
- ☐ UPS

Serial Number:	
----------------	--

2 Verifications:

	Pass	Comments
General status.	☐	
Display status.	☐	
Unbalance detector (use 2 adjointed cards).	☐	
Stop button.	☐	
Electromagnetic lock lid opening.	☐	
Manual lid opening.	☐	
Buzzer beeps when centrifugation is finished.	☐	

3 DG SPIN speed and time for Program 1:

	Measured Speed (rpm)	Range Speed (rpm)	Acceptable
Program 1		980 to 1000	☐

Note: The verification of centrifuge speed implies verification of the equipment time basis and therefore, independent verification of the timer is not necessary.

Reference:	REGD-0008373	Version:	1.0; CURRENT; Most-Recent; Effective	Page	3 of 3
Status:	Effective	Title:	DG SPIN OQ	Effective Date:	05/04/2011

4 ☐ Centrifuging some real cards:

 User / Operator to run a real batch to check overall performance execution.

Operator signature: _____