

RESULTS BY WORKLIST

Serial Number: 00	System SW Version: 3.2.5.9	Date Range: 8/27/2012 12:31:35 PM to 8/27/2012 1:31:34 PM
Assay: CT/GC		
Worklist IDs: Version: 1.38 (For In Vitro Diagnostic Use) ML: 613010 Exp: 11/15/2013		
Operators: marrieta		
Specimen Tested: 60	CT POS: 11 GC POS: 10	
Status Flag Legend:		
Reviewed By: _____ Date: _____ Verified By: _____ Date: _____		

CT/GC*Version: 1.38 (For In Vitro Diagnostic Use)***Worklist ID:**

Time	Rack-Tube #	Sample ID	Total RLU	CT Result	GC Result	Flags
8/27/2012 12:31:35 PM	532-02	CT Positive Control - Lot 608400	1278	CT POS	GC neg	
8/27/2012 12:31:35 PM	532-01	GC Positive Control - Lot 610260	844	CT neg	GC POS	
8/27/2012 12:31:35 PM	664-01	PANEL-A-606029	5	CT neg	GC neg	
8/27/2012 12:31:35 PM	664-02	PANEL-A-606029	7	CT neg	GC neg	
8/27/2012 12:31:35 PM	664-03	PANEL-A-606029	3	CT neg	GC neg	
8/27/2012 12:36:35 PM	664-04	PANEL-A-606029	6	CT neg	GC neg	
8/27/2012 12:36:35 PM	664-05	PANEL-A-606029	3	CT neg	GC neg	
8/27/2012 12:36:35 PM	664-06	PANEL-A-606029	4	CT neg	GC neg	
8/27/2012 12:36:35 PM	664-07	PANEL-A-606029	3	CT neg	GC neg	
8/27/2012 12:36:35 PM	664-08	PANEL-A-606029	2	CT neg	GC neg	
8/27/2012 12:41:35 PM	664-09	PANEL-A-606029	2	CT neg	GC neg	
8/27/2012 12:41:35 PM	664-10	PANEL-A-606029	3	CT neg	GC neg	
8/27/2012 12:41:35 PM	664-11	PANEL-A-606029	4	CT neg	GC neg	
8/27/2012 12:41:35 PM	664-12	PANEL-A-606029	2	CT neg	GC neg	
8/27/2012 12:41:35 PM	664-13	PANEL-A-606029	2	CT neg	GC neg	
8/27/2012 12:46:35 PM	664-14	PANEL-A-606029	4	CT neg	GC neg	

CT/GC**Version: 1.38 (For In Vitro Diagnostic Use)**

Time	Rack-Tube #	Sample ID	Total RLU	CT Result	GC Result	Flags
8/27/2012 12:46:35 PM	664-15	PANEL-A-606029	6	CT neg	GC neg	
8/27/2012 12:46:35 PM	544-01	PANEL-A-606029	3	CT neg	GC neg	
8/27/2012 12:46:35 PM	544-02	PANEL-A-606029	3	CT neg	GC neg	
8/27/2012 12:46:35 PM	544-03	PANEL-B-606397	1218	CT POS	GC neg	
8/27/2012 12:51:35 PM	544-04	PANEL-A-606029	5	CT neg	GC neg	
8/27/2012 12:51:35 PM	544-05	PANEL-A-606029	3	CT neg	GC neg	
8/27/2012 12:51:35 PM	544-06	PANEL-C-606300	875	CT neg	GC POS	
8/27/2012 12:51:35 PM	544-07	PANEL-A-606029	3	CT neg	GC neg	
8/27/2012 12:51:35 PM	544-08	PANEL-A-606029	2	CT neg	GC neg	
8/27/2012 12:56:35 PM	544-09	PANEL-D-613977	2053	CT POS	GC POS	
8/27/2012 12:56:35 PM	544-10	PANEL-A-606029	5	CT neg	GC neg	
8/27/2012 12:56:35 PM	544-11	PANEL-A-606029	8	CT neg	GC neg	
8/27/2012 12:56:35 PM	544-12	PANEL-B-606397	1263	CT POS	GC neg	
8/27/2012 12:56:35 PM	544-13	PANEL-A-606029	3	CT neg	GC neg	
8/27/2012 1:01:35 PM	544-14	PANEL-A-606029	5	CT neg	GC neg	
8/27/2012 1:01:35 PM	544-15	PANEL-D-613977	2279	CT POS	GC POS	
8/27/2012 1:01:35 PM	474-01	PANEL-A-606029	4	CT neg	GC neg	
8/27/2012 1:01:35 PM	474-02	PANEL-A-606029	3	CT neg	GC neg	
8/27/2012 1:01:35 PM	474-03	PANEL-D-613977	2258	CT POS	GC POS	
8/27/2012 1:06:35 PM	474-04	PANEL-A-606029	3	CT neg	GC neg	
8/27/2012 1:06:35 PM	474-05	PANEL-A-606029	3	CT neg	GC neg	
8/27/2012 1:06:35 PM	474-06	PANEL-B-606397	1307	CT POS	GC neg	
8/27/2012 1:06:35 PM	474-07	PANEL-A-606029	4	CT neg	GC neg	
8/27/2012 1:06:35 PM	474-08	PANEL-A-606029	2	CT neg	GC neg	
8/27/2012 1:11:34 PM	474-09	PANEL-C-606300	957	CT neg	GC POS	
8/27/2012 1:11:34 PM	474-10	PANEL-A-606029	6	CT neg	GC neg	
8/27/2012 1:11:34 PM	474-11	PANEL-A-606029	3	CT neg	GC neg	
8/27/2012 1:11:34 PM	474-12	PANEL-D-613977	2325	CT POS	GC POS	
8/27/2012 1:11:34 PM	474-13	PANEL-A-606029	5	CT neg	GC neg	
8/27/2012 1:16:34 PM	474-14	PANEL-A-606029	3	CT neg	GC neg	
8/27/2012 1:16:34 PM	474-15	PANEL-B-606397	1371	CT POS	GC neg	
8/27/2012 1:16:34 PM	441-01	PANEL-A-606029	2	CT neg	GC neg	

CT/GC**Version: 1.38 (For In Vitro Diagnostic Use)**

Time	Rack-Tube #	Sample ID	Total RLU	CT Result	GC Result	Flags
8/27/2012 1:16:34 PM	441-02	PANEL-A-606029	5	CT neg	GC neg	
8/27/2012 1:16:34 PM	441-03	PANEL-C-606300	972	CT neg	GC POS	
8/27/2012 1:21:34 PM	441-04	PANEL-A-606029	3	CT neg	GC neg	
8/27/2012 1:21:34 PM	441-05	PANEL-A-606029	10	CT neg	GC neg	
8/27/2012 1:21:34 PM	441-06	PANEL-D-613977	2344	CT POS	GC POS	
8/27/2012 1:21:34 PM	441-07	PANEL-A-606029	3	CT neg	GC neg	
8/27/2012 1:21:34 PM	441-08	PANEL-A-606029	3	CT neg	GC neg	
8/27/2012 1:26:34 PM	441-09	PANEL-B-606397	1357	CT POS	GC neg	
8/27/2012 1:26:34 PM	441-10	PANEL-A-606029	5	CT neg	GC neg	
8/27/2012 1:26:34 PM	441-11	PANEL-A-606029	6	CT neg	GC neg	
8/27/2012 1:26:34 PM	441-12	PANEL-C-606300	1089	CT neg	GC POS	
8/27/2012 1:26:34 PM	441-13	PANEL-A-606029	4	CT neg	GC neg	
8/27/2012 1:31:34 PM	441-14	PANEL-A-606029	10	CT neg	GC neg	
8/27/2012 1:31:34 PM	441-15	PANEL-D-613977	2134	CT POS	GC POS	