

RESULTS BY WORKLIST

Serial Number: XXXXXXXX	System SW Version: 3.2.2.18	Date Range: 3/12/2012 1:17:56 PM to 3/12/2012 1:17:56 PM
Assay: CT		
Worklist IDs: 000184-20120312-01 Version: 1.32 (For In Vitro Diagnostic Use) ML: 594473 Exp: 10/15/2012		
Operators: XXXXXXXX		
Specimen Tested: 3	Invalid: 3	
Status Flag Legend: ch - Total RLU for Control is Over Maximum Limit. p - Assay Processing Error. x - Invalid Results Due To Control Failure.		
Reviewed By: _____	Date: _____	Verified By: _____ Date: _____

CT

Version: 1.32 (For In Vitro Diagnostic Use)

Worklist ID: 000184-20120312-01

Time	Rack-Tube #	Sample ID	Total RLU	CT Result	Flags
3/12/2012 1:17:56 PM	195-00	GC Positive Control - Lot 602231	65	CT EQUIV	ch
3/12/2012 1:17:56 PM	195-00	CT Positive Control - Lot 602230	3326	CT POS	
3/12/2012 1:17:56 PM	195-00	XXXXXXX	3620	Invalid	x, p
3/12/2012 1:17:56 PM	884-00	XXXXXXX	3764	Invalid	x, p
3/12/2012 1:17:56 PM	884-00	XXXXXXX	3303	Invalid	x, p