

RESULTS BY WORKLIST

Serial Number: 00334	System SW Version: 3.2.4.6	Date Range: 7/16/2012 3:48:27 PM to 7/16/2012 4:08:26 PM
Assay: CT		
Worklist IDs: Version: 1.37 (For In Vitro Diagnostic Use) ML: 605942 Exp: 6/15/2013		
Operators: CHudson		
Specimen Tested: 23	Invalid: 23	
Status Flag Legend: ch - Total RLU for Control is Over Maximum Limit. p - Assay Processing Error. x - Invalid Results Due To Control Failure.		
Reviewed By: _____	Date: _____	Verified By: _____ Date: _____

CT

Version: 1.37 (For In Vitro Diagnostic Use)

Worklist ID:

Time	Rack-Tube #	Sample ID	Total RLU	CT Result	Flags
7/16/2012 3:48:27 PM	341-00	GC Positive Control - Lot 606054	67	CT EQUIV	ch
7/16/2012 3:48:27 PM	341-00	CT Positive Control - Lot 605926	1917	CT POS	
7/16/2012 3:48:27 PM	341-00	14634103	59	Invalid	x, p
7/16/2012 3:48:27 PM	341-00	14651915	1895	Invalid	x, p
7/16/2012 3:48:27 PM	341-00	14652384	1794	Invalid	x, p
7/16/2012 3:53:26 PM	341-00	14652515	5870	Invalid	x, p
7/16/2012 3:53:26 PM	341-00	14653005	481	Invalid	x, p
7/16/2012 3:53:26 PM	341-00	14653714	4635	Invalid	x, p
7/16/2012 3:53:26 PM	341-00	14657894	4248	Invalid	x, p
7/16/2012 3:53:27 PM	341-00	14657978	3237	Invalid	x, p
7/16/2012 3:58:26 PM	341-00	14658007	6750	Invalid	x, p
7/16/2012 3:58:26 PM	341-00	14658281	6883	Invalid	x, p
7/16/2012 3:58:26 PM	341-00	14658320	6358	Invalid	x, p
7/16/2012 3:58:26 PM	341-00	14658366	6893	Invalid	x, p

CT**Version: 1.37 (For In Vitro Diagnostic Use)**

Time	Rack-Tube #	Sample ID	Total RLU	CT Result	Flags
7/16/2012 3:58:26 PM	341-00	14658624	7004	Invalid	x, p
7/16/2012 4:03:26 PM	450-00	14658704	6437	Invalid	x, p
7/16/2012 4:03:26 PM	450-00	14658921	6710	Invalid	x, p
7/16/2012 4:03:26 PM	450-00	14658966	6662	Invalid	x, p
7/16/2012 4:03:26 PM	450-00	14659127	6694	Invalid	x, p
7/16/2012 4:03:26 PM	450-00	14659168	6716	Invalid	x, p
7/16/2012 4:08:25 PM	450-00	14660858	6845	Invalid	x, p
7/16/2012 4:08:26 PM	450-00	14661735	6739	Invalid	x, p
7/16/2012 4:08:26 PM	450-00	14661966	6607	Invalid	x, p
7/16/2012 4:08:26 PM	450-00	14662410	6594	Invalid	x, p
7/16/2012 4:08:26 PM	450-00	14666008	6575	Invalid	x, p