

RESULTS BY WORKLIST

Serial Number:	00334	System SW Version:	3.2.4.6	Date Range:	7/16/2012 4:13:26 PM to 7/16/2012 4:18:26 PM
Assay:	GC				
Worklist IDs:	Version: 1.37 (For In Vitro Diagnostic Use) ML: 605766 Exp: 4/15/2013				
Operators:	CHudson				
Specimen Tested:	7	Invalid:	7		
Status Flag Legend:	ch - Total RLU for Control is Over Maximum Limit. p - Assay Processing Error. x - Invalid Results Due To Control Failure.				
Reviewed By:	_____	Date:	_____	Verified By:	_____
				Date:	_____

GC

Version: 1.37 (For In Vitro Diagnostic Use)

Worklist ID:

Time	Rack-Tube #	Sample ID	Total RLU	GC Result	Flags
7/16/2012 4:13:26 PM	353-00	CT Positive Control - Lot 607368	54	GC EQUIV	ch
7/16/2012 4:13:26 PM	353-00	GC Positive Control - Lot 608399	483	GC POS	
7/16/2012 4:13:26 PM	450-00	14658966	108	Invalid	x, p
7/16/2012 4:13:26 PM	353-00	14658215	674	Invalid	x, p
7/16/2012 4:13:26 PM	353-00	14658224	653	Invalid	x, p
7/16/2012 4:18:25 PM	353-00	14658227	1334	Invalid	x, p
7/16/2012 4:18:25 PM	353-00	14658316	1318	Invalid	x, p
7/16/2012 4:18:25 PM	353-00	14660916	1141	Invalid	x, p
7/16/2012 4:18:26 PM	353-00	14661077	52	Invalid	x, p