

RESULTS BY WORKLIST

Serial Number: XXXXXXXX	System SW Version: 3.2.2.18	Date Range: 2/24/2012 6:32:35 PM to 2/24/2012 7:22:34 PM
Assay: HPV		
Worklist IDs: 00185-20120224-01 Version: 1.50 (For In Vitro Diagnostic Use) ML: 603035 Exp: 2/15/2013 Analyte Cutoff: 0 IC Cutoff: 0		
Operators: XXXXXXXX		
Specimen Tested: 47	Invalid: 47	
Status Flag Legend: f - Calibration failed. M - Sample ID of a new test request was manually entered via the keyboard when a barcode was not readable by the instrument. p - Assay processing error.		
Reviewed By: _____	Date: _____	Verified By: _____ Date: _____

HPV

Version: 1.50 (For In Vitro Diagnostic Use)

Worklist ID: 00185-20120224-01

Time	Rack-Tube #	Sample ID	IC RLU	IC Result	RLU	S/CO	Result	Flags
2/24/2012 6:32:35 PM	459-00	Negative Calibrator - Lot 596865	244662	Invalid	0	0.00	Invalid	p
2/24/2012 6:32:35 PM	459-00	Negative Calibrator - Lot 596865	232855	Invalid	0	0.00	Invalid	p
2/24/2012 6:32:35 PM	459-00	Negative Calibrator - Lot 596865	249208	Invalid	0	0.00	Invalid	p
2/24/2012 6:32:35 PM	459-00	Positive Calibrator - Lot 596867	235356	Invalid	1320697	0.00	Invalid	p
2/24/2012 6:32:35 PM	459-00	Positive Calibrator - Lot 596867	230614	Invalid	1294088	0.00	Invalid	p
2/24/2012 6:37:35 PM	459-00	Positive Calibrator - Lot 596867	260754	Invalid	1326150	0.00	Invalid	p
2/24/2012 6:37:35 PM	459-00	XXXXXXX	484091	Invalid	1903098	0.00	Invalid	M, f, p
2/24/2012 6:37:35 PM	459-00	XXXXXXX	353049	Invalid	854245	0.00	Invalid	M, f, p
2/24/2012 6:37:35 PM	459-00	XXXXXXX	425750	Invalid	1968003	0.00	Invalid	M, f, p
2/24/2012 6:37:35 PM	459-00	XXXXXXX	211468	Invalid	0	0.00	Invalid	M, f, p
2/24/2012 6:42:35 PM	459-00	XXXXXXX	228331	Invalid	1867	0.00	Invalid	M, f, p
2/24/2012 6:42:35 PM	459-00	XXXXXXX	168539	Invalid	0	0.00	Invalid	M, f, p
2/24/2012 6:42:35 PM	459-00	XXXXXXX	456564	Invalid	2330282	0.00	Invalid	M, f, p
2/24/2012 6:42:35 PM	459-00	XXXXXXX	271382	Invalid	1522856	0.00	Invalid	M, f, p

HPV**Version: 1.50 (For In Vitro Diagnostic Use)**

Time	Rack-Tube #	Sample ID	IC RLU	IC Result	RLU	S/CO	Result	Flags
2/24/2012 6:42:35 PM	459-00	XXXXXXXX	223203	Invalid	0	0.00	Invalid	M, f, p
2/24/2012 6:47:35 PM	459-00	XXXXXXXX	204514	Invalid	0	0.00	Invalid	M, f, p
2/24/2012 6:47:35 PM	459-00	XXXXXXXX	237421	Invalid	5700	0.00	Invalid	M, f, p
2/24/2012 6:47:35 PM	459-00	XXXXXXXX	214387	Invalid	0	0.00	Invalid	M, f, p
2/24/2012 6:47:35 PM	459-00	XXXXXXXX	228870	Invalid	0	0.00	Invalid	M, f, p
2/24/2012 6:47:35 PM	682-00	XXXXXXXX	222804	Invalid	1250257	0.00	Invalid	M, f, p
2/24/2012 6:52:35 PM	682-00	XXXXXXXX	263831	Invalid	1263180	0.00	Invalid	M, f, p
2/24/2012 6:52:35 PM	682-00	XXXXXXXX	226837	Invalid	1272892	0.00	Invalid	M, f, p
2/24/2012 6:52:35 PM	682-00	XXXXXXXX	225611	Invalid	0	0.00	Invalid	M, f, p
2/24/2012 6:52:35 PM	682-00	XXXXXXXX	239423	Invalid	0	0.00	Invalid	M, f, p
2/24/2012 6:52:35 PM	682-00	XXXXXXXX	235150	Invalid	0	0.00	Invalid	M, f, p
2/24/2012 6:57:34 PM	682-00	XXXXXXXX	234346	Invalid	0	0.00	Invalid	M, f, p
2/24/2012 6:57:34 PM	682-00	XXXXXXXX	215110	Invalid	0	0.00	Invalid	M, f, p
2/24/2012 6:57:34 PM	682-00	XXXXXXXX	325940	Invalid	1829005	0.00	Invalid	M, f, p
2/24/2012 6:57:34 PM	682-00	XXXXXXXX	224799	Invalid	0	0.00	Invalid	M, f, p
2/24/2012 6:57:35 PM	682-00	XXXXXXXX	233197	Invalid	2705	0.00	Invalid	M, f, p
2/24/2012 7:02:34 PM	682-00	XXXXXXXX	207753	Invalid	0	0.00	Invalid	M, f, p
2/24/2012 7:02:34 PM	682-00	XXXXXXXX	189438	Invalid	4640	0.00	Invalid	M, f, p
2/24/2012 7:02:34 PM	682-00	XXXXXXXX	221783	Invalid	0	0.00	Invalid	M, f, p
2/24/2012 7:02:34 PM	682-00	XXXXXXXX	206762	Invalid	1160241	0.00	Invalid	M, f, p
2/24/2012 7:02:34 PM	683-00	XXXXXXXX	238937	Invalid	0	0.00	Invalid	M, f, p
2/24/2012 7:07:34 PM	683-00	XXXXXXXX	231642	Invalid	0	0.00	Invalid	M, f, p
2/24/2012 7:07:34 PM	683-00	XXXXXXXX	226216	Invalid	0	0.00	Invalid	M, f, p
2/24/2012 7:07:34 PM	683-00	XXXXXXXX	230870	Invalid	0	0.00	Invalid	M, f, p
2/24/2012 7:07:34 PM	683-00	XXXXXXXX	246128	Invalid	0	0.00	Invalid	M, f, p
2/24/2012 7:07:34 PM	683-00	XXXXXXXX	473565	Invalid	1186303	0.00	Invalid	M, f, p
2/24/2012 7:12:34 PM	683-00	XXXXXXXX	211024	Invalid	1184159	0.00	Invalid	M, f, p
2/24/2012 7:12:34 PM	683-00	XXXXXXXX	593040	Invalid	1995474	0.00	Invalid	M, f, p
2/24/2012 7:12:34 PM	683-00	XXXXXXXX	229095	Invalid	893114	0.00	Invalid	M, f, p
2/24/2012 7:12:34 PM	683-00	XXXXXXXX	593621	Invalid	1897763	0.00	Invalid	M, f, p
2/24/2012 7:12:34 PM	683-00	XXXXXXXX	201625	Invalid	0	0.00	Invalid	M, f, p
2/24/2012 7:17:34 PM	683-00	XXXXXXXX	235565	Invalid	0	0.00	Invalid	M, f, p

HPV**Version: 1.50 (For In Vitro Diagnostic Use)**

Time	Rack-Tube #	Sample ID	IC RLU	IC Result	RLU	S/CO	Result	Flags
2/24/2012 7:17:34 PM	683-00	XXXXXXXX	242294	Invalid	1209643	0.00	Invalid	M, f, p
2/24/2012 7:17:34 PM	683-00	XXXXXXXX	444672	Invalid	861440	0.00	Invalid	M, f, p
2/24/2012 7:17:34 PM	683-00	XXXXXXXX	212360	Invalid	0	0.00	Invalid	M, f, p
2/24/2012 7:17:34 PM	456-00	XXXXXXXX	226551	Invalid	1271288	0.00	Invalid	M, f, p
2/24/2012 7:22:33 PM	456-00	XXXXXXXX	230707	Invalid	0	0.00	Invalid	M, f, p
2/24/2012 7:22:34 PM	456-00	XXXXXXXX	667667	Invalid	1899427	0.00	Invalid	M, f, p
2/24/2012 7:22:34 PM	456-00	XXXXXXXX	230658	Invalid	0	0.00	Invalid	M, f, p

For Internal Use Only