

RESULTS BY WORKLIST

Serial Number:	XXXXXXX	System SW Version:	3.2.2.18	Date Range:	11/11/2011 12:12:09 PM to 11/11/2011 12:27:09 PM
Assay:	HPV				
Worklist IDs:	00001-20111111-01 Version: 1.47 (For In Vitro Diagnostic Use) ML: 597942 Exp: 8/15/2012 Analyte Cutoff: 0 IC Cutoff: 0				
Operators:	XXXXXXX				
Specimen Tested:	10	Invalid: 10			
Status Flag Legend:	ag - Results interpretation error. f - Calibration failed. g - Results interpretation error. M - Sample ID of a new test request was manually entered via the keyboard when a barcode was not readable by the instrument. p - Assay processing error.				
Reviewed By:	_____	Date:	_____	Verified By:	_____
				Date:	_____

HPV

Version: 1.47 (For In Vitro Diagnostic Use)

Worklist ID: 00001-20111111-01

Time	Rack-Tube #	Sample ID	IC RLU	IC Result	RLU	S/CO	Result	Flags
11/11/2011 12:12:09 PM	561-00	Negative Calibrator - Lot 588262	132861	Invalid	108054	0.00	Invalid	p, g, ag
11/11/2011 12:12:09 PM	561-00	Negative Calibrator - Lot 588262	141525	Invalid	97609	0.00	Invalid	p, g, ag
11/11/2011 12:12:09 PM	561-00	Negative Calibrator - Lot 588262	140164	Invalid	90733	0.00	Invalid	p, g, ag
11/11/2011 12:12:09 PM	561-00	Positive Calibrator - Lot 588263	298960	Invalid	447628	0.00	Invalid	p, ag
11/11/2011 12:12:09 PM	561-00	Positive Calibrator - Lot 588263	341753	Invalid	452369	0.00	Invalid	p, ag
11/11/2011 12:17:10 PM	561-00	Positive Calibrator - Lot 588263	374327	Invalid	588159	0.00	Invalid	p
11/11/2011 12:17:10 PM	561-00	XXXXXXXX	577919	Invalid	1092510	0.00	Invalid	M, f, p
11/11/2011 12:17:10 PM	561-00	XXXXXXXX	310072	Invalid	459713	0.00	Invalid	M, f, p
11/11/2011 12:17:10 PM	561-00	XXXXXXXX	402928	Invalid	831373	0.00	Invalid	M, f, p
11/11/2011 12:17:10 PM	561-00	XXXXXXXX	135391	Invalid	70938	0.00	Invalid	M, f, p
11/11/2011 12:22:09 PM	561-00	XXXXXXXX	239935	Invalid	0	0.00	Invalid	M, f, p
11/11/2011 12:22:09 PM	561-00	XXXXXXXX	278458	Invalid	1137911	0.00	Invalid	M, f, p

HPV**Version: 1.47 (For In Vitro Diagnostic Use)**

Time	Rack-Tube #	Sample ID	IC RLU	IC Result	RLU	S/CO	Result	Flags
11/11/2011 12:22:10 PM	561-00	XXXXXXXX	231712	Invalid	1265487	0.00	Invalid	M, f, p
11/11/2011 12:22:10 PM	561-00	XXXXXXXX	225973	Invalid	1130283	0.00	Invalid	M, f, p
11/11/2011 12:22:10 PM	561-00	XXXXXXXX	211481	Invalid	0	0.00	Invalid	M, f, p
11/11/2011 12:27:09 PM	561-00	XXXXXXXX	236113	Invalid	0	0.00	Invalid	M, f, p

For Internal Use Only