

RESULTS BY WORKLIST

Serial Number:	XXXXXXX	System SW Version:	3.2.2.18	Date Range:	10/19/2011 9:30:15 PM to 10/19/2011 10:10:14 PM
Assay:	HPV				
Worklist IDs:	00001-20111019-01 Version: 1.47 (For In Vitro Diagnostic Use) ML: 594238 Exp: 11/15/2011 Analyte Cutoff: 0 IC Cutoff: 0				
Operators:	XXXXXXX				
Specimen Tested:	31	Invalid: 31			
Status Flag Legend:	ag - Results interpretation error. f - Calibration failed. g - Results interpretation error. M - Sample ID of a new test request was manually entered via the keyboard when a barcode was not readable by the instrument. p - Assay processing error.				
Reviewed By:	_____	Date:	_____	Verified By:	_____
Date:	_____				

HPV

Version: 1.47 (For In Vitro Diagnostic Use)

Worklist ID: 00001-20111019-01

Time	Rack-Tube #	Sample ID	IC RLU	IC Result	RLU	S/CO	Result	Flags
10/19/2011 9:30:15 PM	682-00	Negative Calibrator - Lot 588262	106742	Invalid	177406	0.00	Invalid	p, g, ag
10/19/2011 9:30:15 PM	682-00	Negative Calibrator - Lot 588262	106305	Invalid	177144	0.00	Invalid	p, g, ag
10/19/2011 9:30:15 PM	682-00	Negative Calibrator - Lot 588262	100326	Invalid	175781	0.00	Invalid	p, g, ag
10/19/2011 9:30:15 PM	682-00	Positive Calibrator - Lot 588263	223639	Invalid	402575	0.00	Invalid	p, ag
10/19/2011 9:30:15 PM	682-00	Positive Calibrator - Lot 588263	222965	Invalid	385589	0.00	Invalid	p, ag
10/19/2011 9:35:15 PM	682-00	Positive Calibrator - Lot 588263	203264	Invalid	394971	0.00	Invalid	p, ag
10/19/2011 9:35:16 PM	682-00	XXXXXXXX	240435	Invalid	463405	0.00	Invalid	M, f, p
10/19/2011 9:35:16 PM	682-00	XXXXXXXX	86210	Invalid	161892	0.00	Invalid	M, f, p
10/19/2011 9:35:16 PM	682-00	XXXXXXXX	91428	Invalid	158946	0.00	Invalid	M, f, p
10/19/2011 9:35:16 PM	682-00	XXXXXXXX	242528	Invalid	641558	0.00	Invalid	M, f, p
10/19/2011 9:40:15 PM	682-00	XXXXXXXX	123403	Invalid	261944	0.00	Invalid	M, f, p
10/19/2011 9:40:15 PM	682-00	XXXXXXXX	112014	Invalid	128014	0.00	Invalid	M, f, p

HPV**Version: 1.47 (For In Vitro Diagnostic Use)**

Time	Rack-Tube #	Sample ID	IC RLU	IC Result	RLU	S/CO	Result	Flags
10/19/2011 9:40:15 PM	682-00	XXXXXXXX	131078	Invalid	126213	0.00	Invalid	M, f, p
10/19/2011 9:40:15 PM	682-00	XXXXXXXX	125656	Invalid	125298	0.00	Invalid	M, f, p
10/19/2011 9:40:15 PM	682-00	XXXXXXXX	286274	Invalid	525056	0.00	Invalid	M, f, p
10/19/2011 9:50:15 PM	682-00	XXXXXXXX	368138	Invalid	730492	0.00	Invalid	M, f, p
10/19/2011 9:50:15 PM	682-00	XXXXXXXX	178150	Invalid	61741	0.00	Invalid	M, f, p
10/19/2011 9:50:15 PM	682-00	XXXXXXXX	73848	Invalid	61315	0.00	Invalid	M, f, p
10/19/2011 9:50:15 PM	682-00	XXXXXXXX	294815	Invalid	739040	0.00	Invalid	M, f, p
10/19/2011 9:50:15 PM	441-00	XXXXXXXX	113943	Invalid	69634	0.00	Invalid	M, f, p
10/19/2011 9:55:15 PM	441-00	XXXXXXXX	243899	Invalid	0	0.00	Invalid	M, f, p
10/19/2011 9:55:15 PM	441-00	XXXXXXXX	475428	Invalid	1885486	0.00	Invalid	M, f, p
10/19/2011 9:55:15 PM	441-00	XXXXXXXX	249413	Invalid	0	0.00	Invalid	M, f, p
10/19/2011 9:55:15 PM	441-00	XXXXXXXX	241331	Invalid	0	0.00	Invalid	M, f, p
10/19/2011 9:55:15 PM	441-00	XXXXXXXX	197534	Invalid	987929	0.00	Invalid	M, f, p
10/19/2011 10:00:14 PM	441-00	XXXXXXXX	240375	Invalid	0	0.00	Invalid	M, f, p
10/19/2011 10:00:14 PM	441-00	XXXXXXXX	219130	Invalid	0	0.00	Invalid	M, f, p
10/19/2011 10:00:14 PM	441-00	XXXXXXXX	159275	Invalid	0	0.00	Invalid	M, f, p
10/19/2011 10:00:14 PM	441-00	XXXXXXXX	231807	Invalid	0	0.00	Invalid	M, f, p
10/19/2011 10:00:14 PM	441-00	XXXXXXXX	226872	Invalid	7733	0.00	Invalid	M, f, p
10/19/2011 10:05:14 PM	441-00	XXXXXXXX	419305	Invalid	1200931	0.00	Invalid	M, f, p
10/19/2011 10:05:14 PM	441-00	XXXXXXXX	217086	Invalid	0	0.00	Invalid	M, f, p
10/19/2011 10:05:14 PM	441-00	XXXXXXXX	226321	Invalid	1231930	0.00	Invalid	M, f, p
10/19/2011 10:05:14 PM	441-00	XXXXXXXX	269265	Invalid	637945	0.00	Invalid	M, f, p
10/19/2011 10:05:14 PM	456-00	XXXXXXXX	186868	Invalid	1149584	0.00	Invalid	M, f, p
10/19/2011 10:10:14 PM	456-00	XXXXXXXX	233141	Invalid	0	0.00	Invalid	M, f, p
10/19/2011 10:10:14 PM	456-00	XXXXXXXX	235780	Invalid	0	0.00	Invalid	M, f, p