

Laminar Flow Maintenance Checklist

Report Number: _____

Company			
Address			
Owner/Contact		Rm Number:	
Telephone/ Fax		Rm Name:	
Manufacturer/model		Type:	
Serial Number		Unit Number:	
Maintenance Date			
Next maintenance			
Calibrated by			

General Cleaning				
No.	Item	Description	Completed (Y,N,or NA)	Remarks
1	Inner Work Tray	Wipe off the inner work tray using IPA, suitable cleaning agent or disinfectant		
2	Inner Walls	Wipe off the inner rear and side walls using IPA, suitable cleaning agent or disinfectant		
3	Air Diffuser	Wipe off the aluminium air diffuser using IPA, suitable cleaning agent or disinfectant		
4	Stainless Steel Arm Rest	Wipe off the stainless armrest using IPA, suitable cleaning agent or disinfectant		
5	Exterior	Visually inspect the exterior and clean the exterior of the cabinet		

Preventive Maintenance				
No.	Item	Description	Completed (Y,N,or NA)	Remarks
1	Test Label	Check the test label, ensuring that it is within certification period.		
2	Power Supply	Check that the power supply is suitable and safely connected.		
3	Exhaust	Check the exhaust outlet/filter and airflow sensor are not obstructed. Tighten airflow sensor if needed.		
4	Operations	Check the operation of the cabinet. a. Control panel and LED indicators b. Inner Fan c. Fluorescent Lamps d. UV Lamp, if any e. Cabinet Electrical Outlet/Socket, if any		
5	Airflow	Check the LCD for the inflow and downflow readings which are to be within manufacturer's accepted range. Adjust to balance airflow if needed. This can only be done by qualified technician/engineer.		
6	Electrical Module/ Panel	Electrical Module Assembly is to be serviced by qualified engineer.		

		a. Check for any traces of burnt marks b. Check on the fuses. c. Tighten all electrical connections		
7	Bottle trap (If any)	Visual Check of bottle trap for any leakage/ chokes		

Equipment Used			
Equipment Description	Application	Serial	Calibration Due
TSI	Wind Velocity Measurement		
FLUKE MULTIMETER	Voltage, Current & Resistance		