

GRIFOLS Diagnostic Grifols, S.A.	Reference:	REGD-0000749		
	Version:	17.0; Most-Recent; Effective; CURRENT		
	Status:	Effective	Purpose:	Uncontrolled Copy
Class:	Control del Diseño		Effective date:	24-Apr-2018
Title:	SM DG57. WADiana Service Manual Index - April 2018 Edition			

DIAGNOSTIC GRIFOLS, S.A.

DG-57 **WADiana Service Manual Index** **April 2018 Edition**

SERVICE MANUAL

DIAGNOSTIC GRIFOLS, S.A.
Passeig Fluvial 24
08150 Parets del Vallès
Barcelona, SPAIN

GRIFOLS Diagnostic Grifols, S.A.	Reference:	REGD-0000749		
	Version:	17.0; Most-Recent; Effective; CURRENT		
	Status:	Effective	Purpose:	Uncontrolled Copy
Class:	Control del Diseño		Effective date:	24-Apr-2018
Title:	SM DG57. WADiana Service Manual Index - April 2018 Edition			

Confidential Information. It is not allowed the total or partial reproduction of this document without prior authorization.

Print date:	24-Apr-2018	Page	3 of 9
--------------------	-------------	-------------	--------

GRIFOLS Diagnostic Grifols, S.A.	Reference:	REGD-0000749		
	Version:	17.0; Most-Recent; Effective; CURRENT		
	Status:	Effective	Purpose:	Uncontrolled Copy
Class:	Control del Diseño		Effective date:	24-Apr-2018
Title:	SM DG57. WADiana Service Manual Index - April 2018 Edition			

Note: Technical information in REGD-xxxxxxx_vx documents start from the second page onwards.

Symbols used

This manual uses the following symbols to facilitate the location of specific sections.

	DANGER! Any kind of injury –critical or not– for people, that only can be diminished following the instructions shown.
	WARNING! Any kind of injury for people for which protection systems are provided, but it is convenient the QUALIFIED TECHNICIAN knows it.
	CAUTION! Any kind of harmful situation for equipment or others. This situation can be diminished using this indication or with other protections.
	NOTE: Used for clarification and complementary or emphasised information.

Invention patents

This instrument is protected by international patents applicable to the entire machine or parts thereof.

The manufacturers reserve the right to take legal action in order to protect their interests.

This manual should be used in conjunction with the WADiana Compact User's Manual and the Diana System Software Manual, especially with the separate WADiana C software Manual.

The information found in this document may be changed without prior notification.

GRIFOLS Diagnostic Grifols, S.A.	Reference:	REGD-0000749		
	Version:	17.0; Most-Recent; Effective; CURRENT		
	Status:	Effective	Purpose:	Uncontrolled Copy
Class:	Control del Diseño		Effective date:	24-Apr-2018
Title:	SM DG57. WADiana Service Manual Index - April 2018 Edition			

CONTENTS:

DESCRIPTION	DOCUMENT TITLE	DOCUMENT NUMBER
General information		
1. Functional Characteristics	Instrument General Overview	REGD-0000884_v5
2. Technical specifications	Technical specifications	REGD-0000774_v5
Setting up and starting the equipment		
3. Installation guidelines	Installation guidelines	REGD-0000784_v3
4. Unpacking and location	Unpacking and location remarks	REGD-0000775_v3
5. Software installation	Software installation	REGD-0000777_v6
	Software uninstallation	REGD-0000778_v5
	Windows Embedded Standard 7 Multi-language Installation	REGD-0016476_v1
	Disable/Enable 'C:' Hard Disk Unit software protection	REGD-0010049_v2
6. Installation qualification	Image acquisition board installation	REGD-0000779_v4
	Image acquisition board uninstallation	REGD-0000780_v7
	Installation qualification	REGD-0007336_v6
7. Operation qualification	Operation qualification	REGD-0007337_v8
8. Networking	Networking	REGD-0000781_v3
	LIS Connection	REGD-0000782_v9
9. Software configuration	Software configuration	REGD-0000783_v6
10. Locked System	Closed OS setup	REGD-0000845_v5
11. Software structure	Files and folders structure	REGD-0000838_v4
Preventive Maintenance		
12. User maintenance	User maintenance	REGD-0000785_v4
13. Preventive Maintenance	Preventive Maintenance	REGD-0000786_v10
14. Decontamination Procedure	Decontamination procedure	REGD-0000787_v6
15. Lubrication	Lubrication points and procedures	REGD-0000841_v4
16. EMC	EMC Considerations	REGD-0000848_v6
Troubleshooting		
17. Troubleshooting	Troubleshooting	REGD-0000788_v3
18. Error codes	Error codes list	REGD-0000789_v8
19. Initialitizing	Initialisation troubleshooting	REGD-0000837_v3
20. Log file	Log file description	REGD-0000847_v3
Adjusting and checking the equipment		
21. Technical service software	Technical Service Software overview	REGD-0000844_v8
22. WADDiagnostic	WADDiagnostic	REGD-0000846_v7
23. Power supplies and communications	Serial number configuration	REGD-0000790_v4
	Switched power supply 24V adjustment and verification	REGD-0000791_v4
	8V and 12V supplies verification	REGD-0000792_v3
	Power fuse status verification	REGD-0000793_v3
24. Sample -Reagent carousel module	Pca samples-reagents-2 control current limits adjustment	REGD-0000794_v3
	Home sensors adjustment	REGD-0000795_v4
	Reagent pipetting position offset adjustment	REGD-0000796_v3
	Samples-Reagent carousel module verification	REGD-0000797_v4
	PCA Samples-Reagents-2 and	REGD-0000798_v3
	PCA Centrifuge-2 fuse status verification	

Confidential Information. It is not allowed the total or partial reproduction of this document without prior authorization.

Print date:	24-Apr-2018	Page	5 of 9
--------------------	-------------	-------------	--------

GRIFOLS Diagnostic Grifols, S.A.	Reference:	REGD-0000749		
	Version:	17.0; Most-Recent; Effective; CURRENT		
	Status:	Effective	Purpose:	Uncontrolled Copy
Class:	Control del Diseño		Effective date:	24-Apr-2018
Title:	SM DG57. WADiana Service Manual Index - April 2018 Edition			

DESCRIPTION	DOCUMENT TITLE	DOCUMENT NUMBER
25. Block incubator module	Temperature adjustment and verification	REGD-0000799_v3
	Block incubator module fuse status verification	REGD-0000800_v3
26. Centrifugal module	Rotor vertical adjustment	REGD-0000801_v3
	PCA Centrifuge-2 control current limits adjustment	REGD-0000802_v3
	Unbalance detector adjustment and verification	REGD-0000803_v4
	Centrifuge module verification	REGD-0000804_v3
	PCA Samples-Reagents-2 and	REGD-0000798_v3
	PCA Centrifuge-2 fuse status verification	
27. Robot module and level detection	PCA Robot 4 axis control current limits adjustment	REGD-0000806_v4
	Gripper position and probe height adjustment in incubator	REGD-0000807_v5
	Gripper adjustment in centrifuge position	REGD-0000808_v4
	Probe X-Y fine adjustment in pipetting positions	REGD-0001221_v2
	Robot module verification	REGD-0000805_v5
	PCA Robot-2 4 axis fuse status verification	REGD-0000809_v3
28. Cameras module	Samples camera adjustment	REGD-0000810_v5
	Reagents camera adjustment	REGD-0000811_v4
	Reader camera adjustment	REGD-0000812_v8
	Reference Image creation	REGD-0000814_v5
	Cameras module verification	REGD-0000815_v4
29. Fluid module	Pressure Regulator Adjustment	REGD-0000816_v4
	Pressure Detector Adjustment	REGD-0000817_v4
	Vacuum Detector Adjustment	REGD-0000818_v4
	Load Cells adjustment	REGD-0000819_v4
	Clot detector sensitivity adjustment	REGD-0000820_v3
	Fluid module verification	REGD-0000821_v5
30. Miscellaneous	General door adjustment	REGD-0000849_v4
Replacement Procedures		
31. Tubing	Tubing replacement	REGD-0000822_v5
32. Cabling	Cabling replacement	REGD-0000823_v5
33. Motors	Motors replacement	REGD-0000824_v4
34. PCA	PCA's replacement	REGD-0000825_v6
35. Gripper/pins	Gripper and gripper pins replacement	REGD-0000826_v4
36. Peltier cells	Peltier cells replacement	REGD-0000843_v6
37. Leadscrew and ball bearings	Leadscrew and ball bearings replacement	REGD-0000842_v4
38. Cameras	Cameras replacement	REGD-0000827_v7
39. Probe	Probe replacement	REGD-0000828_v7
40. Pumps	Pumps replacement	REGD-0000829_v5
41. Electrovalves	Electrovalves replacement	REGD-0000830_v4
42. Digital dilutor / Syringe	Syringe / Dilutor replacement	REGD-0000831_v5
Verification Procedures		
43. Centrifuge verification procedure	Centrifuge verification procedure	REGD-0000832_v5
44. Pipetting verification procedure	Pipetting system verification procedure	REGD-0000833_v6
45. Incubator verification procedure	Incubator user verification procedure	REGD-0000834_v3
46. Dispensing verification procedure	"Dispense verification" user verification procedure	REGD-0000835_v4

Confidential Information. It is not allowed the total or partial reproduction of this document without prior authorization.

Print date:	24-Apr-2018	Page	6 of 9
--------------------	-------------	-------------	--------

GRIFOLS Diagnostic Grifols, S.A.	Reference:	REGD-0000749		
	Version:	17.0; Most-Recent; Effective; CURRENT		
	Status:	Effective	Purpose:	Uncontrolled Copy
Class:	Control del Diseño		Effective date:	24-Apr-2018
Title:	SM DG57. WADiana Service Manual Index - April 2018 Edition			

DESCRIPTION	DOCUMENT TITLE	DOCUMENT NUMBER
-------------	----------------	-----------------

Compatibility with instruments with s/n < 057-118-xxxx

47. Unbalance sensor adjustment	Unbalance detector adjustment and verification	REGD-0000803_v4
48. Fluid module	Pressure Detector Adjustment	REGD-0000817_v4
	Vacuum Detector Adjustment	REGD-0000818_v4
49. Sample -Reagent carousel module	Home sensors adjustment	REGD-0000795_v4

GRIFOLS Diagnostic Grifols, S.A.	Reference:	REGD-0000749		
	Version:	17.0; Most-Recent; Effective; CURRENT		
	Status:	Effective	Purpose:	Uncontrolled Copy
Class:	Control del Diseño		Effective date:	24-Apr-2018
Title:	SM DG57. WADiana Service Manual Index - April 2018 Edition			

DOCUMENT CHANGE HISTORY

Document change history			
Version	Date	Change	Rationale
1.0	2005/02/17	Document REGD-0000749_v1 replaces document 57000V0M002.	Uploaded document to GxPharma.
2.0	2005/04/18	Updated document numbers with the latest version	Updated content due to changes in documents version.
3.0	2005/05/05	Updated document numbers with the latest version	Updated content due to changes in documents version.
4.0	2005/05/06	Updated document numbers with the latest version	Updated content due to changes in documents version.
5.0	2005/05/09	Updated document numbers with the latest version	Updated content due to changes in documents version.
6.0	2005/09/21	Updated document numbers with the latest version	Updated content due to changes in documents version.
7.0	2005/12/12	Updated document numbers with the latest version	Updated content due to changes in documents version.
8.0	2011/05/06	Added the Document Change History to the document.	Adoption of ISO 8601 date format to avoid ambiguity.
8.0	2011/05/06	Updated document numbers with the latest version	Updated content due to changes in documents version.
8.0	2011/05/06	ID-GELSTATION and PROVUE erased from the title.	Service manual intended for Grifols market.
8.0	2011/05/06	Added document REGD-0000779 "Image adquisition board" installation.	Added document REGD-0000779 that was not present in the content of the software installation.
9.0	2011/08/17	Updated document numbers with the latest version	Updated content due to changes in documents version.
9.0	2011/08/17	Added documents REGD-0007336 and REGD-0007337.	Include installation qualification and operation qualification documents.
9.0	2011/08/17	Document REGD-0000826 title has change.	Backslash replaced by words.
9.0	2011/08/17	Document title.	Commercial reference substituted by project reference.
9.0	2011/08/17	Document header.	Increase information to easy document search.
10.0	2011/09/30	Versions updated.	Documents updated.
11.0	2011/10/13	Versions updated.	Document REGD-0000786_v7 replaced by REGD-0000786_v8.
11.0	2011/10/13	Document title.	Document title has change.
12.0	2012/10/30	Documents updated.	Following documents have been updated: -REGD-0000782 -REGD-0000797 -REGD-0007337
13.0	2013/03/25	Document updated.	REGD-0000812 has been updated.
14.0	October 2013	Document updated	Following documents have been updated: -REGD-0000782 -REGD-0000812
15.0	December 2014	Document updated	-Updated information in several documents -New Matrox 4Sight GP computer

Confidential Information. It is not allowed the total or partial reproduction of this document without prior authorization.

Print date:	24-Apr-2018	Page	8 of 9
--------------------	-------------	-------------	--------

GRIFOLS Diagnostic Grifols, S.A.	Reference:	REGD-0000749		
	Version:	17.0; Most-Recent; Effective; CURRENT		
	Status:	Effective	Purpose:	Uncontrolled Copy
Class:	Control del Diseño		Effective date:	24-Apr-2018
Title:	SM DG57. WADiana Service Manual Index - April 2018 Edition			
16.0	April 2016	Document updated	Updated information in several documents.	
17.0	April 2018	Updated document versions	Include main changes related to Diana System software version 4.4.0.	