

GRIFOLS Diagnostic Grifols, S.A.	Reference:	REGD-0000784	Version:	3.0, CURRENT
	Status:	Effective	Effective Date:	19-May-2011
Title:	SM DG57. Installation guidelines.			Page 2 of 4

DG-57 Service Manual

INSTALLATION GUIDELINES

Procedure description
This document provides some guidelines to follow to properly install the instrument. It is structured in a list of items, relating them to the appropriate document where full description of action points can be found. <i>Note: Document REGD-0000784_v3 replaces document REGD-0000784_v2.</i>

GRIFOLS Diagnostic Grifols, S.A.	Reference:	REGD-0000784	Version:	3.0, CURRENT
	Status:	Effective	Effective Date:	19-May-2011
Title:	SM DG57. Installation guidelines.			Page 3 of 4

EQUIPMENT

SET UP

1. ---

PROCEDURE

Process	Expected Results
1. Unpacking and instrument location: see document REGD-0000775.	---
2. Sol. A and Sol. B bottles filling: fill Sol A. and Sol B. bottles following the instructions for use of both products.	---
3. Balancing card insertion in first balancing card position.	---
4. Image Acquisition Board (Matrox) installation: see document REGD-0000779.	---
5. Computer and mains connection: - Computer and UPS requirements: see document REGD-0000774 - Connection to mains supply: Plug the instrument to UPS or mains inlet. - Connection to computer: connect RS-232 and video cables from PC to instrument.	---
6. Software installation: see document REGD-0000777. Configuration of working directory and language.	---
7. Software configuration: - Serial port configuration: see document REGD-0000844 to change COM port configuration if not connected to COM1 port (default). - Serial number configuration: see document REGD-0000790. - Registry entries configuration: see document REGD-0000783. - Locked system: see document REGD-0000845. - Reference image: see document REGD-0000814.	---
8. User software configuration (user configuration is normally performed by product specialists, refer to software manual): - Test files (GRU). - Templates. - User Configuration options.	---
9. Initial tests: - Modular tests: see documents REGD-0000797, REGD-0000799, REGD-0000804, REGD-0000805, REGD-0000815, and REGD-0000821. - Waddiagnostic: see document REGD-0000846. - Real batch test: run a real batch to check overall performance.	---
10. Networking: see document REGD-0000781.	---
11. Host connection: see document REGD-0000782.	---

GRIFOLS Diagnostic Grifols, S.A.	Reference:	REGD-0000784	Version:	3.0, CURRENT
	Status:	Effective	Effective Date:	19-May-2011
Title:	SM DG57. Installation guidelines.			Page 4 of 4

DOCUMENT CHANGE HISTORY

Document change history			
Version	Date	Change	Rationale
2.0	2005/05/09	<i>Document REGD-0000784_v2 replaces document 57011V0M000.</i>	Document's reference changed.
3.0	See document header	Replaced commercial name by project name.	Delete commercial name.
3.0	See document header	<i>Document REGD-0000784_v3 replaces document REGD-0000784_v2.</i>	Document version changed.
3.0	See document header	Header title has changed.	Previous header title has been substituted by project and document title.