

GRIFOLS Diagnostic Grifols, S.A.	Reference:	REGD-0000801	Version:	3.0, CURRENT
	Status:	Effective	Effective Date:	16-Jun-2011
Title:	SM DG57-DG517. Rotor vertical adjustment.			Page 2 of 4

DG-57 / DG-517 Service Manual

ROTOR VERTICAL ADJUSTMENT

Procedure description
This procedure describes how adjust the rotor height with the height adjustment tool. <i>Note: Document REGD-0000801_v3 replaces document REGD-0000801_v2.</i>

GRIFOLS Diagnostic Grifols, S.A.	Reference:	REGD-0000801	Version:	3.0, CURRENT
	Status:	Effective	Effective Date:	16-Jun-2011
Title:	SM DG57-DG517. Rotor vertical adjustment.			Page 3 of 4


EQUIPMENT

1. Centrifuge Rotor Height Adjustment tool (P/N: 232084).
2. Centrifuge Rotor Key (P/N: 232022).

SET UP


1. Remove the centrifuge cover.
2. Remove all centrifuge card holders.
3. Extract the Rotor and the nut.

PROCEDURE

Process	Expected Results
1. Insert the Centrifuge Rotor Height Adjustment tool in the nut (see picture 1) with the magnet of the tool taking the nut.  Picture 1	---
2. Insert the rotor in the instrument and put the set of the nut with the tool inside the rotor axis.  Picture 2	---
3. Raise the Rotor until the Rotor touches the nut.  Picture 3	---

GRIFOLS Diagnostic Grifols, S.A.	Reference:	REGD-0000801	Version:	3.0, CURRENT
	Status:	Effective	Effective Date:	16-Jun-2011
Title:	SM DG57-DG517. Rotor vertical adjustment.			Page 4 of 4

4. **Holding the rotor** touching the nut, **press down the tool** until the tool leans on the motor axis.



Picture 4

5. Holding the position of all the pieces **screw the nut** to the centrifuge motor axis using the Rotor key. ---
6. Replace all centrifuge card holders and replace the centrifuge cover. ---
7. Adjust the gripper in the centrifuge following the procedure REGD-0000808. ---

DOCUMENT CHANGE HISTORY

Document change history			
Version	Date	Change	Rationale
2.0	2005/05/09	Document REGD-0000801_v2 replaces document 57029VOM000.	Document reference changed.
3.0	See document header.	Added previous version 'data', 'change' and 'rationale' inside document change history section.	Update change history for previous document history modification.
3.0	See document header.	Document REGD-0000801_v3 replaces document REGD-0000801_v2.	Document version changed.
3.0	See document header.	Document title.	Commercial reference substituted by project reference.
3.0	See document header.	Document header.	Increase information to easy document search.