

GRIFOLS Diagnostic Grifols, S.A.	Reference:	REGD-0000818	Version:	4.0, CURRENT
	Status:	Effective	Effective Date:	08-Jul-2011
Title:	SM DG57-DG517. Vacuum detector adjustment.			Page 2 of 5

DG-57 / DG-517 Service Manual

VACUUM DETECTOR ADJUSTMENT

Procedure description
This procedure describes how adjust the value of the potentiometer P1 of PCA Control IO57-2 in order to detect when there is enough vacuum in the fluidic system. Note: Document REGD-0000818_v4 replaces document REGD-0000818_v3.

GRIFOLS Diagnostic Grifols, S.A.	Reference:	REGD-0000818	Version:	4.0, CURRENT
	Status:	Effective	Effective Date:	08-Jul-2011
Title:	SM DG57-DG517. Vacuum detector adjustment.			Page 3 of 5

EQUIPMENT

1. Manometer + racor type CPC (1 unit code 232201 and 1 unit 231860).
2. Multimeter.
3. Calculator.



NOTE:

Ground in all measurements is TP23

SET UP

1. Technical Service software started.
2. Execute the ST software (Technical Service software).
3. Enter the Technical Service password (d+g_diana).
4. Pumps turned off.



CAUTION!

Short-circuit between any powered part and ground, may damage the instrument.

PROCEDURE

Process	Expected Results
1. Measure and record the value in TP11* [mV]. Picture 1	---

GRIFOLS Diagnostic Grifols, S.A.	Reference:	REGD-0000818	Version:	4.0, CURRENT
	Status:	Effective	Effective Date:	08-Jul-2011
Title:	SM DG57-DG517. Vacuum detector adjustment.			Page 4 of 5

Process	Expected Results
2. Disconnect the racor n°24 from the racor n°39 (bottle) and connect the manometer between these two racors. 3. Activate the pumps: Access I/O tab, click on Pumps checkbox. 4. Wait until the manometer measure is stable. 5. Measure and record the value of the vacuum [mbar] 6. Measure and record the value in TP11* [mV] again. See picture 1. 7. Calculate: $\text{Vacuum} = \frac{\text{Value (step 6[mV])} - \text{Value (Step 1[mV])}}{\text{Value (step 5[mBar])}} \times 305 + \text{Value (step 1[mV])}$ 8. Place the multimeter in TP10. 9. Adjust potentiometer P1 until voltage measured in TP10 is Vacuum (result of step 7).	--- Pumps turn on Below -330 mbar --- --- --- --- ---

* Note: instruments with s/n < 057-118-xxxx, be aware that corresponding value should be measured at TP9 instead of TP11.

GRIFOLS Diagnostic Grifols, S.A.	Reference:	REGD-0000818	Version:	4.0, CURRENT
	Status:	Effective	Effective Date:	08-Jul-2011
Title:	SM DG57-DG517. Vacuum detector adjustment.			Page 5 of 5

DOCUMENT CHANGE HISTORY

Document change history			
Version	Date	Change	Rationale
3.0	2005/05/06	<i>Document REGD-0000818_v3 replaces document 57045V0M001.</i>	Document reference changed.
3.0	2005/05/06	-Included vacuum test points names. -I/O Board version dependence. - modified expected value in step 4.	Update procedure.
4.0	See document header	Added 'previous version 'data', 'change' and 'rationale' inside document change history section.	Update change history for previous document history modification.
4.0	See document header.	<i>Document REGD-0000818_v4 replaces document REGD-0000818_v3.</i>	Document version changed.
4.0	See document header.	Equipment section updated.	Added spare part codes.
4.0	See document header.	Document title.	Commercial reference substituted by project reference.
4.0	See document header.	Document header.	Increase information to easy document search.