

GRIFOLS Diagnostic Grifols, S.A.	Reference:	REGD-0007336		
	Version:	6.0; Most-Recent; Effective; CURRENT		
	Status:	Effective	Purpose:	Uncontrolled Copy
Class:	Control del Diseño		Effective date:	19-Dec-2014
Title:	SM DG57. Installation qualification.			

DG57 Service Manual

INSTALLATION QUALIFICATION

Procedure description
During the Installation Qualification (IQ) phase, all of the activities associated with properly installing the instrument at the users' site are documented. The IQ procedure should be executed during the installation of the instrument at customer site. <i>Note: Document REGD-0007336_v6 replaces document REGD-0007336_v5.</i>

GRIFOLS Diagnostic Grifols, S.A.	Reference:	REGD-0007336		
	Version:	6.0; Most-Recent; Effective; CURRENT		
	Status:	Effective	Purpose:	Uncontrolled Copy
Class:	Control del Diseño		Effective date:	19-Dec-2014
Title:	SM DG57. Installation qualification.			

INSTALLATION QUALIFICATION (IQ)

Installation Qualification REGD-0007336 is related to DG-57 service manual (REGD-0000749).

Instrument:	DG57	Form version:	6.0
Serial number:		Code:	
Installation date:			

Customer Site:		Customer Name:	
Department:			
Tel:		Fax:	
		E-Mail:	

Service Engineer Name:			
Department:			
Tel:		Fax:	
		E-Mail:	

Action	Done
Review with the operator temperature/humidity/altitude requirements indicated at Service Manual (Technical specifications attached).	☐

Procedure	Expected result (with acceptance criteria)	Result obtained	Pass/Fail
Visual Inspection.			
Check that no defects are visible.	No defects are visible		
Any part or component which is separated from its original position.	No parts neither components are separated		
Any impairment which reduce the safety of the instrument.	Non impairment are detected		
Any impairment which affect to performance of the instrument.	Non impairment are detected		
Installation requirements.			
Place for installation. (Wadiana)	Width > 1000mm Depth > 600mm Height (with door closed)> 650mm Weight > 87kg		
Place for installation (Computer)	The necessary space to install the computer connected to Wadiana Compact is available.		

GRIFOLS Diagnostic Grifols, S.A.	Reference:	REGD-0007336		
	Version:	6.0; Most-Recent; Effective; CURRENT		
	Status:	Effective	Purpose:	Uncontrolled Copy
Class:	Control del Diseño		Effective date:	19-Dec-2014
Title:	SM DG57. Installation qualification.			

Place for installation.	Space of 10-20 cm for service on the right and left sides of the instrument is recommended.		
	No direct sunlight reaches the instrument.		
Inclination. (can be adjusted by means of its feet).	Instrument installed in a horizontal position.		
The instrument should not be used adjacent to or stacked with other instrument.	No adjacent or stacked with other instrument		
Do not block air openings nor place the instrument on a soft surface which might block them.	Do not block air openings		
Mains connection must have a ground terminal.	Ground terminal is present		
Mains supply maximum voltage fluctuations.	±10% of nominal voltage.		
Packing list			
Check that the packing list is present and that all the parts included are present.	Packing list and all the parts included are present		
Documentation			
Check that the Certificate of Analysis (CoA) is present.	CoA present		

Computer / peripherals/communications	Available	Comments
Computer.	☐	
Printer.	☐	
Communication with LIS:		
Serial communication	☐	
File communication	☐	
Others.	☐	

Comments:

Service Engineer signature:

Customer signature:

GRIFOLS Diagnostic Grifols, S.A.	Reference:	REGD-0007336		
	Version:	6.0; Most-Recent; Effective; CURRENT		
	Status:	Effective	Purpose:	Uncontrolled Copy
Class:	Control del Diseño		Effective date:	19-Dec-2014
Title:	SM DG57. Installation qualification.			

DOCUMENT CHANGE HISTORY

Document change history			
Version	Date	Change	Rationale
4.0	2011/05/23	Replaced commercial name by project name.	Delete commercial name.
4.0	2011/05/23	Note: Document REGD-0001336_v4 replaces document REGD-0001336_v3.	Document version changed.
4.0	2011/05/23	Added 'procedure description' section.	Unify service manual documents style.
4.0	2011/05/23	Updated 'action' section explanation.	Current 'technical specifications' version will be attached.
4.0	2011/05/23	Header title has changed.	Previous header title has been substituted by project and document title.
5.0	September 2011	Added service manual index references.	Document related to a service manual.
5.0	September 2011	Added number of pages	Number of pages in the page bottom side.
5.0	September 2011	Erased recommendation to perform OQ after IQ.	IQ not related to OQ.
6.0	December 2014	The document will only apply to DG57 project	A new IQ document is created for DG517 project
6.0	December 2014	Added Place for computer installation.	Not included in previous versions
6.0	December 2014	Change in 'Documentation' section	Certificate of Analysis substitutes Certificate of Conformity

Confidential Information. It is not allowed the total or partial reproduction of this document without prior authorization.

Print date:	19-Dec-2014	Page	5 of 5
--------------------	-------------	-------------	--------