

GRIFOLS Diagnostic Grifols, S.A.	Reference:	REGD-0016476		
	Version:	1.0; CURRENT; Most-Recent; Effective		
	Status:	Effective	Purpose:	Uncontrolled Copy
Class:	Genérico Regulado		Effective date:	19-Dec-2014
Title:	SM DG57. Windows Embedded Standard 7 Multi-language Installation			

DG-57 Service Manual

WINDOWS EMBEDDED STANDARD 7 MULTI-LANGUAGE INSTALLATION

Procedure description
This procedure describes how to install Windows 7 (embedded) image in a Matrox 4Sight PC. Installation of the image is performed in the following situations: • New hard disk installation. • Reinstallation of a corrupted Operating System. • Performing an Operating System upgrade Note: The installation of Windows Embedded Standard 7 O.S. does not include the installation of Diana System software. Refer to service manual document REGD-0000777 for details on the latter procedure.

GRIFOLS Diagnostic Grifols, S.A.	Reference:	REGD-0016476		
	Version:	1.0; CURRENT; Most-Recent; Effective		
	Status:	Effective	Purpose:	Uncontrolled Copy
Class:	Genérico Regulado		Effective date:	19-Dec-2014
Title:	SM DG57. Windows Embedded Standard 7 Multi-language Installation			

EQUIPMENT

1.	DG-57 instrument.
2.	PC/laptop running under Windows 7 O.S.
3.	Computer Matrox 4sight GP
4.	USB drive (NTFS format) with at least 8 GB capacity
5.	WES7 Matrox Multi-Language Install Disc.

SET UP

1.	Instrument wiring connected.
----	------------------------------

Step 1. Preparing the installation device

1.	Insert the USB drive in a PC/laptop running under Windows 7	--
2.	Open a DOS command line by clicking the Start button and typing 'cmd' into the search box, and then pressing Enter	DOS window will open
3.	Type the following commands diskpart list disk select disk X Note: 'X' is the USB drive disk number. Before continuing, make sure that no other unit is being selected by mistake, as all its contents will be erased during the following process clean create partition primary select partition 1 active format fs=ntfs quick assign exit	The USB drive is empty (ntfs format)
4.	Finally copy the contents of the DVD containing the WES7 Matrox Multi-Language image into the USB drive	The USB drive contains the required files for the O.S installation


Step 2. Booting from the USB memory unit connected to Matrox 4Sight GP

Process	Expected Results
5. With the PC Matrox 4Sight GP switched off: - Connect a keyboard, mouse and screen - Connect the USB drive containing the Windows Embedded Standard 7 Multi-Language installation files.	--
6. Switch on the PC Matrox 4Sight GP and access to the BIOS by pressing ESC .	The BIOS menu is accessed

GRIFOLS Diagnostic Grifols, S.A.	Reference:	REGD-0016476		
	Version:	1.0; CURRENT; Most-Recent; Effective		
	Status:	Effective	Purpose:	Uncontrolled Copy
Class:	Genérico Regulado		Effective date:	19-Dec-2014
Title:	SM DG57. Windows Embedded Standard 7 Multi-language Installation			

Process	Expected Results
7. Once in the Bios menu: 1. Select 'Boot Manager' from the menu options (Picture 1) 2. Select the USB memory unit to boot from it	The system boots from the USB memory unit.
 Picture 1	

Step 3. Installing Windows Embedded Standard 7 Multi-Language from USB memory stick

Process	Expected Results
8. Matrox installation wizard (ResQDisk Utility) will be started (Picture 2)	License terms window appears
 Picture 2 Select 'Next'	
9. The license Terms for Microsoft Windows installation will appear. Accept the terms by selecting 'I agree'	Next installation window appears.
10. Choose the 'Restore OS' action	Next installation window appears.
11. In the next dialogue, select 'Factory image 32 bits' and choose the 4Sight GP hard disk from the drop-down list. Then click on 'Go' to start the installation procedure	Next installation window appears.


GRIFOLS Diagnostic Grifols, S.A.	Reference:	REGD-0016476		
	Version:	1.0; CURRENT; Most-Recent; Effective		
	Status:	Effective	Purpose:	Uncontrolled Copy
Class:	Genérico Regulado		Effective date:	19-Dec-2014
Title:	SM DG57. Windows Embedded Standard 7 Multi-language Installation			

Process	Expected Results
12. An alert message will be prompted informing that the operation about to start will reformat and erase all data on the selected hard disk. Make sure that all the valuable data stored in the system (if any) has been backed up, and select 'Yes' to start the installation procedure.	Installation procedure is started
13. After some time, a window indicating that the process has finished successfully will be prompted. Click OK and the system will reboot. Note: Before starting again, the USB drive must be unplugged	At this point the PC will reboot.
14. The system will start with a Set Up window allowing the user to choose among several languages (Picture 3). Select the desired one and click 'Next' Picture 3	Next configuration window appears (Picture 4)
15. Select the corresponding country / region, time / currency and keyboard layout and click 'Next' Picture 4	Next configuration window appears (Picture 5)

GRIFOLS Diagnostic Grifols, S.A.	Reference:	REGD-0016476		
	Version:	1.0; CURRENT; Most-Recent; Effective		
	Status:	Effective	Purpose:	Uncontrolled Copy
Class:	Genérico Regulado		Effective date:	19-Dec-2014
Title:	SM DG57. Windows Embedded Standard 7 Multi-language Installation			

Process	Expected Results
16. Introduce user name and click 'Next' Picture 5	Next configuration window appears
17. Type a password (if any) and click 'Next'	Next configuration window appears
18. Accept Windows' license terms and click 'Next'	Next configuration window appears (Picture 6)
19. Select 'Use recommended settings' Picture 6	Next configuration window appears (Picture 7)

GRIFOLS Diagnostic Grifols, S.A.	Reference:	REGD-0016476		
	Version:	1.0; CURRENT; Most-Recent; Effective		
	Status:	Effective	Purpose:	Uncontrolled Copy
Class:	Genérico Regulado		Effective date:	19-Dec-2014
Title:	SM DG57. Windows Embedded Standard 7 Multi-language Installation			

Process	Expected Results
20. Select time zone and date and click 'Next' to finish the installation procedure  Picture 7	The system restarts X times automatically
21. Once the installation process is finished the PC is ready to be used.	--
22. At this point, the C drive will be unprotected against writing. Install Diana system software following the 'Software installation' document (REGD-0000777) and then protect the system following the 'Disable/Enable 'C:' hard disk unit software protection' document (REGD-0010049).	--

GRIFOLS Diagnostic Grifols, S.A.	Reference:	REGD-0016476		
	Version:	1.0; CURRENT; Most-Recent; Effective		
	Status:	Effective	Purpose:	Uncontrolled Copy
Class:	Genérico Regulado		Effective date:	19-Dec-2014
Title:	SM DG57. Windows Embedded Standard 7 Multi-language Installation			

DOCUMENT CHANGE HISTORY

Document change history			
Version	Date	Change	Rationale
1.0	December 2014	NA	First version of the document.