Installation Qualification Procedure

© Grifols — DG-AT-TSM-000165 1.12.0

Device

Instrument	Serial Number	Code	Installation Date
Erytra Eflexis	EF-99021	EFX	2026-10-29

Customer

Name	Site	Department	Telephone
KFSH	Jeddah		

Email

Service Engineer

Name	Department	Telephone	Email
M. Aldweik			

Action

Checked/done	Action	Comments
☑	Review with the operator temperature/humidity/altitude requirements indicated in the Erytra Eflexis Instruction for Use, technical specification section.	

Visual Inspection

Action	Results	Comments
Check that there are no visible defects.	FAIL	scratch on cover
There is no part or component detached from its original position.	PASS	
There are no safety related anomalies.	PASS	
There are no performance related anomalies.	FAIL	

Installation Requirements

Action	Results	Comments
Check that the place for installation matches the following requirements: • Width > 1100 mm • Depth > 710 mm • Height > 910 mm • Weight > 173 kg	PASS	
The instrument should not be used adjacent to or stacked with another instrument.	PASS	
Instrument must be placed in a non inclined surface.	PASS	
Do not block air openings nor place the instrument on a soft surface which might block them.	PASS	

Mains connection must have a ground terminal.	PASS	
Nominal voltage shown in the instrument matches with the supply voltage.	PASS	

Installation Levelling

Action	Results	Comments
Perform the levelling verification according to the procedure to make sure the calculated reference point is within the allowed range.	PASS	

Packing List

Action	Results	Comments
Check that packing list is present and that all the parts included are present.	PASS	

Documentation

Action	Results	Comments
Check that the certificate of Analysis (CoA) is present. If that's not the case, check for its availability in the Grifols Technical Library.	PASS	
Check if the IFU provided with the instrument installed is the current version.	PASS	

Peripherals

Item	Comments
☐ Printer	
☑ External barcode reader	
☑ Others	
☐ RS232 connection	
☐ Network connection	
☐ USB connection	
☐ Fluidic drain	

Date

2026-10-29

Service engineer signature

M. Aldweik

Customer signature

Dr. A

© Grifols — DG-AT-TSM-000165 1.12.0